


Analytical intercomparison of heavy metals in precipitation, 2009 and 2010

Hilde Thelle Uggerud and Anne-Gunn Hjellbrekke



NILU : EMEP/CCC-Report 4/2011
REFERENCE : O-7726
DATE : SEPTEMBER 2011

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**Analytical intercomparison of heavy metals in
precipitation, 2009 and 2010**

Hilde Thelle Uggerud and Anne-Gunn Hjellbrekke



NILU – Norwegian Institute for Air Research
PO Box 100, NO-2027 Kjeller, Norway

Contents

	Page
1. Analytical intercomparison of heavy metals in precipitation, 2009 and 2010	5
1.1 Introduction	5
1.2 Organization of the intercomparison	5
1.2.1 Organization of the twelfth intercomparison	5
1.2.2 Organization of the thirteenth intercomparison	5
1.3 Intercomparison samples	6
1.4 Data handling.....	6
1.4.1 Data analysis	6
1.4.2 Youden plot	6
1.5 Summary.....	7
2. References	8
Appendix 1 Tables and figures, 2009.....	9
Appendix 2 Tables and figures, 2010.....	29

Analytical intercomparison of heavy metals in precipitation, 2009 and 2010

1. Analytical intercomparison of heavy metals in precipitation, 2009 and 2010

1.1 Introduction

Heavy metals were included in the EMEP's monitoring programme in 1999. 20 countries are reporting data to the heavy metal database. Since EMEP's measurement programme is based on individual national networks, different sampling and analytical methods are applied by the participating laboratories. In order to ensure data comparability, interlaboratory tests are organized by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research. So far 13 intercomparisons have been arranged (Berg and Semb, 1995; Berg and Aas, 2000; Uggerud and Skjelmoen, 2001, 2002, 2003; Uggerud and Hjellbrekke, 2005, 2007, 2008, 2009).

The present report presents results from the twelfth and thirteenth analytical intercomparison of heavy metals in precipitation, which was carried out in 2009 and 2010, respectively. Seven heavy metals were included: Pb, Cd, Cu, Zn, As, Cr, and Ni.

1.2 Organization of the intercomparison

1.2.1 Organization of the twelfth intercomparison

The samples for the twelfth intercomparison were prepared and distributed to 50 laboratories in July 2009.

A total of 43 laboratories, 19 from the EMEP network, reported results within the end of December 2009. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. All tables and figures are given in Appendix 1. Tables 2a and 2b give the names of the participating laboratories together with the number used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Table 3.

1.2.2 Organization of the thirteenth intercomparison

The samples for the thirteenth intercomparison were prepared and distributed to 55 laboratories in July 2010.

A total of 37 laboratories, 17 from the EMEP network, reported results within the end of November 2010. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. All tables and figures are given in Appendix 2. Tables 12a and 12b give the names of the participating laboratories together with the number used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Table 13.

1.3 Intercomparison samples

The four synthetic precipitation samples distributed were made from multi-element standards traceable to NIST-standards. The multi-element standards were conserved with 2.5% HNO₃. The distributed synthetic precipitation samples contained Pb, Cd, Cu, Zn, As, Cr, and Ni in 0.5% HNO₃. Sample H1 and H2 contained concentrations similar to what is normally found in Southern Scandinavia. Sample H3 and H4 contained the elements in concentrations normally found in Central Europe.

All equipments in contact with the samples were soaked in 3% HNO₃ for 4 days. Preparation of the intercomparison samples was carried out in a clean room area.

1.4 Data handling

The data reported for the twelfth intercomparison from the participants are presented in Tables 5-11 and Figures 1-7. An overview of all results is presented in Table 4.

The data reported for the thirteenth intercomparison from the participants are presented in Tables 15-21 and Figures 8-14. An overview of all results is presented in Table 14.

1.4.1 Data analysis

The reported values are presented in the tables in decreasing order together with the number of the laboratory. The expected (theoretical) value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation in percent are also given. After the first statistical run with all results included, the calculation was repeated with the outliers excluded. The outliers (unused) are defined as the results more than two standard deviations from the mean value in the first run.

In EMEP2010, all reported results for Ni in sample H1 showed a systematic error of 30%. This is highly unusual and most probably the theoretical value is wrong. Thus, the expected value for Ni in sample H1 is in all calculations replaced with the median value of the reported values.

1.4.2 Youden plot

Youden plot is a graphical technique, which allows for analysing interlaboratory data, where two samples of equal or similar concentrations have been analysed. The Youden plot visualises systematic errors as well as random errors.

The precipitation samples are made in pairs with similar concentrations and the reported value for one sample is plotted on the x-axis and the reported value of the other sample is plotted on the y-axis. Thus, each point in the plot is representing a pair of results from a single laboratory. Two fully drawn lines represent the expected values of the two samples. Two dotted lines represent the arithmetic mean values in the second statistical run. The lines divide the plot in four

quadrants. A 45°-reference line may be drawn through the intercept of the lines representing the expected values.

If the errors are due to random factors, the points will be evenly distributed around the mean value and be situated in all four quadrants of the chart.

If the errors are due to systematic factors, the results will be close to the 45°-reference line, but situated in the upper right or lower left quadrant.

Ellipses with radii corresponding to the data quality objectives within EMEP are drawn in each plot (see table 1). The data points are colour coded as given in Table 1. Drawn arrows indicate points outside the plot area.

Table 1: Youden-plot parameters.

Radii = DQO	Concentration
25% accuracy or better	Pb, Ni, Cr, As <1 µg/l, Cd <0.5 µg/l, Zn < 10 µg/l, Cu <2 µg/l
15% accuracy or better	Pb, Ni, Cr, As >1 µg/l, Cd >0.5 µg/l, Zn >10 µg/l, Cu >2 µg/l
Criteria	Colour
Within 0.5*DQO	Blue
Within DQO	Green
Within 2*DQO	Orange
> 2*DQO	Red

The length of the perpendicular from an individual point and to the reference line gives a measure of the random error. The perpendicular intercepts the 45°-reference line at a distance from the origin of the fully drawn lines. This distance is a measure of the systematic error.

Youden plots are presented in Figures 1-7 and Figures 8-14 for EMEP2009 and EMEP2010, respectively.

1.5 Summary

The data quality objectives (DQO) in EMEP, states that the accuracy in the laboratory should be better than 15% and 25% for high and low concentration samples, respectively. More than 94% of the results reported in 2009 lies within DQO. In 2010, 90% of the reported results were acceptable.

As in earlier intercomparisons, outliers are defined as values that deviates more than two standard deviations from the mean value. Outliers occur for all samples and all parameters in both intercomparisons. Out of a total of 972 single results reported to EMEP2009, 47 are defined as outliers. This is about 5% of the reported data, which is comparable to earlier intercomparisons. Out of a total of 930 single results reported to emep2010, 38 are defined as outliers. This is about 4% of the reported data, which is slightly better than earlier intercomparisons.

2. References

- Berg, T., Aas, W. (2000) Analytical intercomparison of heavy metals in precipitation 1999. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2000).
- Berg, T., Semb, A. (1995) Preliminary results from HELCOM-EMEP-PARCOM-AMAP analytical intercomparison of heavy metals in precipitation. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Note 1/95).
- Uggerud, H.T., Skjelmoen, J.E. (2001) Analytical intercomparison of heavy metals in precipitation 2000. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 12/2001).
- Uggerud, H.T., Skjelmoen, J.E. (2002) Analytical intercomparison of heavy metals in precipitation 2001. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/2002).
- Uggerud, H.T., Skjelmoen, J.E. (2003) Analytical intercomparison of heavy metals in precipitation 2002. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 7/2003).
- Uggerud, H.T., Hjellbrekke, A.-G. (2005) Analytical intercomparison of heavy metals in precipitation, 2003 and 2004. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 7/2005).
- Uggerud, H.T., Hjellbrekke, A.-G. (2007) Analytical intercomparison of heavy metals in precipitation, 2005 and 2006. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 4/2007).
- Uggerud, H.T., Hjellbrekke, A.-G. (2008) Analytical intercomparison of heavy metals in precipitation, 2007. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2008).
- Uggerud, H.T., Hjellbrekke, A.-G. (2009) Analytical intercomparison of heavy metals in precipitation, 2008. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 5/2009).

Appendix 1

Tables and figures, 2009

Table 2a: Participating laboratories in the EMEP network, 2009. The numbers in front are used in tables.

No	Laboratory identification
1	Federal Environmental Agency, Austria
2	Flemish Environment Agency, Belgium
3	Czech Hydrometeorological Institute, Czech Republic
4	National Environmental Research Institute. Air Pollution Laboratory, Denmark
5	Finnish Meteorological Institute, Finland
6	SGS Multilab, France
8	Umweltbundesamt, Germany
10	Hungarian Meteorological Service, Hungary
13	C.N.R. Istituto Inquinamento Atmosferico, Italy
14	RIVM Laboratory of Inorganic Analytical Chemistry, The Netherlands
15	The Norwegian Institute for Air Research, Norway
16	Inst. Of Meteorology and Water Management, Poland
23	AEA Technology, National Environmental Techn. Centre, United Kingdom
31	Slovak Hydrometeorological Institute, Slovakia
33	Environmental Pollution Observ. Centre, Latvia
36	Hydrometeorological Institute of Slovenia, Slovenia
38	Estonian Environmental Research Centre, Estonia
39	Environmental Monitoring Laboratory, Institute of Environmental Protection, Poland
45	Environmental Impact Monitoring Center, Yerevan, Armenia

Table 2b: Participating laboratories outside the EMEP network, 2009. The number in front of the names is used in tables and figures.

No	Laboratory identification
108	Institut f. Bodenkunde und Standortlehre der TU Dresden, Germany
109	Institut f. Bodenkunde und Waldernährung der Universität, Germany
110	Thüringer Landesanstalt für Landwirtschaft (TTL), Germany
112	Niedersächsische Forstliche Versuchsanstalt (NVF), Germany
113	Fachhochschule Eberswalde, Germany
114	C.N.R. Istituto Italiano di Idrobiologia, Italy
115	Bayerische Landesanstalt f. Wald- und Forstwirtschaft, Germany
118	Forstliche Versuchs- und Forschungsanstalt Baden-Württemberg, Germany
120	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA), Germany
121	Landesamt für Natur und Umwelt, Germany
125	Bayerisches Landesamt für Umweltschutz, Germany
127	Dept. Of Chemistry, Environmental Health Division, Malaysia
129	Ecole Nationale d'ingénieurs de Sfax, Tunisia
132	Comision Chilena De Energia Nuclear, Chile
141	Pollutants Chemical Analysis Centre, Marine Division, Japan
146	Cellule de Research en Environment et Biotechnologies Public Research Center-Gabriel Lippman, Luxembourg
159	CARSO, France
160	Coillte, Wicklow, Ireland
161	National Institute of Chemistry, Slovenia
168	Universite de Bretagne Occidentale, France
169	Lancaster Environment Centre, Centre for Ecology &Hydrology, UK
171	Ecole de Mines de Douai, Departement Chimie et Environment, France
174	Background Monitoring Department, Institute of Global Climate and Ecology(IGCE), Russian Federation
179	Murmansk Environmental Monitoring Center, Russian Federation

Table 3: Analytical techniques used at the participating laboratories for the different elements, 2009.

Lab no	Network	As	Cd	Cr	Cu	Pb	Ni	Zn
1	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
2	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
3	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
4	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	
5	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
6	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
8	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
10	EMEP		GF-AAS			GF-AAS		
13	EMEP		AS voltammetry		AS voltammetry	AS voltammetry		
14	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
15	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
16	EMEP	As methods	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
23	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
31	EMEP	HG-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS
33	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
36	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
38	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS
39	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS
45	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
108	EMEP			ICP-AES	ICP-AES		ICP-AES	ICP-AES
109	EMEP		ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
110	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
112	ICP-Forest		ICP-AES	ICP-MS	ICP-AES	ICP-AES	ICP-AES	ICP-AES
113	Other	ICP- AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
114	ICP-Forest	ICP- AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
115	ICP-Forest	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
118	EMEP	ICP- AES	GF-AAS	ICP-AES	ICP-AES	GF-AAS	ICP-AES	ICP-AES
120	EMEP							ICP-AES
121	EMEP	Hydrid- AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS ICP- OES (H3, H4)	GF-AAS	ICP-AES
125	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
127	EANET	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
129	EMEP		GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS
132	EMEP		ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
141	WMO-GAW		GF-AAS					
146	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
159	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
160	EMEP	As methods	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
161	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
168	Other	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
169	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
171	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
174	WMO-GAW		GAAS (H1 & H2 done with pre-concentrating by evaporation)					
179	WMO-GAW	GAAS	GAAS	GAAS	GF-AAS	GF-AAS	GAAS	F-AAS

Table 4: Reported results for metal determination in precipitation samples, expressed as % deviation from expected value, 2009.

Lab no.	Arsenic				Cadmium				Chromium				Copper				Lead				Nickel				Zinc				
	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	
1	-3	-5	-7	-7	-7	-1	-4	-4	-5	-8	-5	-5	5	-3	-4	-5	0	1	0	1	6	4	-5	-7	7	2	0	1	
2	1	1	4	-2	-4	0	3	-1	1	-3	8	7	-22	-10	8	5	-10	-7	-3	-5	-1	5	4	3	9	5	8	6	
3	-4	-8	-4	-3	-6	-10	-4	-1	10	7	4	7	-5	-3	0	1	-5	-7	0	1	-11	-13	-2	-1	4	1	3	4	
4	< 1.5	< 1.5	-2	-1	< 0.35	< 0.35	-13	-17	< 1.35	< 1.35	-8	-10	< 2.5	< 2.5	-3	-16	< 3.0	< 3.0	-4	-5	< 3.0	< 3.0	-9	-8	6	0	-2	-3	
5	-10	-13	-8	-8	-4	-7	-4	-4	-4	-10	-1	-2	4	2	2	3	2	3	1	2	-7	-8	1	1	6	0	-2	-3	
6	4	4	4	2	< 0.1	< 0.1	-1	-3	-21	-27	-8	-9	-14	-9	-7	-8	-25	-28	3	3	-19	-16	-8	-8	23	19	10	10	
8	-3	-4	-3	-5	4	2	-3	-3	-1	-3	-2	-3	-1	-5	-3	-5	-3	-4	-6	-7	-3	-4	-3	-4	3	0	-1	-3	
10					-16	-33	-1	-2									35	5	-14	-19					18	15	0	-2	
13					-14	17	-6	18					75	57	-15	6	20	10	-4	1									
14	-14	1	-9	-4	-1	13	-4	-4	1	2	-1	-1	-1	-7	-8	-9	-6	7	0	0	-15	1	-4	-3	70	5	-2	-1	
15	-9	-10	0	2	0	-2	-6	0	-18	-17	-7	-3	-19	-24	1	-1	-7	0	0	0	-19	-15	-2	-7	1	-4	-5	-4	
16					0	0	0	0	0	0	0	0	0	0	0	-6	0	0	0	0	-14	-13	0	0	8	-6	-3	-2	
23	-57	20	-8	-1	329	< 0.08	-14	0	25	67	-33	-6	150	43	-9	-9	33	-10	-8	-5	186	-38	-4	-11	233	-13	-7	0	
31	-3	-4	-6	-4	1	-2	-4	-7	15	12	3	3	-1	12	-1	-2	-1	2	0	-3	20	9	2	2	17	13	-1	0	
33	-1	4	0	0	0	0	0	-2	1	-2	2	3	0	0	2	3	0	-10	0	0	14	0	1	1	33	0	3	3	
36	-1	-8	0	0	-10	0	0	-2	1	-5	-4	-6	4	-5	-5	-5	2	-3	-6	-7	-11	-14	-4	-4	15	6	0	0	
38	-4	-2	-3	-3	-14	867	-9	-5	6	7	5	6	7	7	6	6	-13	-8	-10	-6	-6	-4	-1	-1	-12	-11	-14	-10	
39	0	0	0	0	0	0	0	0	0	0	0	0	-6	-14	-6	0	0	30	-8	-7	0	0	0	0	0	0	0	0	
45	-21	-33	8	18	-66	-72	-9	2	-25	-32	4	10	-58	-28	0	10	-25	-24	-13	-8	-38	-30	2	10	11	15	20	26	
108					< 1.35	< 1.35	-18	-7	575	300	55	56	179	161	20	17	< 1.0	< 1.0	-13	-15	< 2.35	< 2.35	-1	8	32	18	-1	1	
109					< 0.1	< 0.1	-29	-67	-38	13	-6	-10	179	161	20	17	< 1.0	< 1.0	-13	-15	< 0.1	< 0.1	-15	-15	< 20	< 20	-4	-4	
110	-23	-10	-18	-19	-14	0	-21	13	-35	-47	-26	-22	793	116	-8	-6	10	91	-11	11	-10	-15	-29	-8	29	7	10	9	
112					10	12	1	2	4	3	3	2	-1	-4	-5	-6	-5	-1	-2	-3	-9	-5	2	1	5	2	0	-1	
113	< 4	< 4	-4	15	< 0.4	< 0.4	11	13	199	306	49	44	< 4	< 4	-34	-20	< 7	< 7	-9	-4	< 1.5	< 1.5	-34	-30	< 5	-31	-9	-8	
114					43	67	-14	-17	-13	-17	-15	-14	-13	-21	-15	-14	-13	-10	-18	-17	-14	-25	-17	-15	2	-11	-15	-16	
115	27	47	9	5	4	-3	2	3	-6	-18	-9	-6	5	-2	-2	-2	3	2	2	2	-3	-4	-90	-4	12	7	5	6	
118	71	200	3	-7	-71	-100	-4	-7	-25	-50	-8	-7	-13	-14	-6	-7	-60	-100	-1	1	29	38	1	-1	-33	-35	-3	-6	
120																									3	-4	-3	-2	
121	4	12	-5	-9	29	0	3	5	23	2	10	13	-25	-23	-9	-13	-3	5	2	1	-29	-25	-2	-5	2	5	4	4	
125	10	9	8	10	3	3	3	5	7	5	-2	-2	6	1	-1	-1	3	6	-1	0	-2	-3	-2	1	15	8	9	9	
127	0	0	3	1	< 0.2	< 0.2	14	0	0	0	2	-1	0	0	4	2	0	0	6	3	< 2	< 2	4	2	8	6	6	4	
129					< 4	< 4	< 4	< 4	4250	5033	350	279	350	900	78	-33	433	< 41	-74	21	< 88	< 88	< 88	< 88	145	1450	16	-77	
132					< 5	< 5	< 5	< 5	11	41	-1	1	< 1	< 1	-26	-27	< 5	< 5	5	9	< 5	< 5	7	-11	-13	-12	-3	-3	
141					-11	-25	-15	-17																					
146	-18	-8	4	12	14	2	3	3	6	8	-3	0	5	-5	-5	-6	2	2	-2	0	10	6	1	3	5	-13	9	8	
159	< 1	< 1	5	3	< 1	< 1	< 1	< 1	< 1	< 1	-28	-29	< 1	< 1	-26	-27	< 1	< 1	-8	-10	< 1	< 1	-24	-28	< 1	< 1	-46	-41	
160																									39	31	15	45	
161	-1	2	1	6	0	0	0	3	2	3	2	4	4	1	-1	-1	3	3	-2	2	-6	-6	-4	-4	7	4	-2	-2	
168	-11	-10	-7	-10	-14	-17	0	0	-24	-20	-12	-10	-16	-23	-18	-17	-1	5	-6	-9	31	9	-7	-3	-3	-8	-3	-5	
169	-11	-11	-5	-8	-9	-13	-8	-10	-6	-5	-6	-9	-5	-8	-8	-10	-7	-8	-8	-10	-9	-10	-8	-11	6	1	-3	-5	
171	1	0	-4	-4	-1	-6	-3	-2	-2	-6	-5	-3	1	-3	-4	-4	-6	-7	-7	-7	-3	-5	-3	-3	8	5	-2	-2	
174					34	12	-27	-22					69	-11	-20	-31	-70	-71	-54	-65									
179	10	-10	-6	-9	0	-17	-19	-17	-8	-12	-11	-12	-3	-24	-6	-4	10	15	-3	5	-14	-6	-25	-26	42	22	1	0	

between ± 25-50%
more than ± 50%
for low theoretical values ((Pb, Ni, Cr, As <1 µg/l, Cd <0.5 µg/l, Cu <2 µg/l)

between ± 15-30%
more than ± 30%
for high theoretical values ((Pb, Ni, Cr, As >1 µg/l, Cd >0.5 µg/l, Cu >2 µg/l)

Table 5: Analytical results for Cr in synthetic precipitation samples, 2009.

Chromium				Chromium			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.800		Theoretical value:		0.600	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		34		Number of laboratories:		34	
Arithmetic mean value:		1.820		Arithmetic mean value:		1.523	
Median:		0.800		Median:		0.595	
Standard deviation		5.835		Standard deviation		5.184	
Rel. st. deviation (%)		320.644		Rel. st. deviation (%)		340.493	
Run 2:				Run 2:			
Number of laboratories:		33		Number of laboratories:		33	
Arithmetic mean value:		0.820		Arithmetic mean value:		0.635	
Median:		0.800		Median:		0.590	
Standard deviation		0.305		Standard deviation		0.347	
Rel. st. deviation (%)		37.194		Rel. st. deviation (%)		54.578	
Results in decreasing order:				Results in decreasing order:			
129	34.800 (*)	39	0.800	129	30.800 (*)	33	0.590
4	< 2.0			4	< 2.0		
108	1.35			108	< 1.34		
159	< 1			159	< 1		
113	2.391	127	0.800	113	2.433	2	0.584
23	1.000	8	0.790	23	1.000	8	0.580
121	0.980	171	0.783	132	0.845	36	0.570
31	0.920	5	0.771	109	0.680	169	0.569
132	0.887	1	0.760	31	0.670	171	0.562
3	0.877	115	0.755	146	0.648	1	0.554
125	0.857	169	0.754	3	0.642	5	0.541
38	0.850	179	0.740	38	0.640	179	0.530
146	0.845	114	0.700	125	0.630	15	0.500
112	0.833	15	0.660	112	0.620	114	0.500
161	0.820	6	0.630	161	0.620	115	0.494
14	0.811	168	0.610	14	0.610	168	0.480
33	0.810	45	0.603	121	0.610	6	0.440
36	0.810	118	0.600	16	0.600	45	0.407
2	0.807	110	0.520	39	0.600	110	0.320
16	0.800	109	0.500	127	0.600	118	0.300
Chromium				Chromium			
Sample no.: H3				Sample no.: H4			
Theoretical value:		6.000		Theoretical value:		7.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		38		Number of laboratories:		38	
Arithmetic mean value:		6.319		Arithmetic mean value:		7.364	
Median:		5.852		Median:		6.838	
Standard deviation		3.532		Standard deviation		3.283	
Rel. st. deviation (%)		55.903		Rel. st. deviation (%)		44.581	
Run 2:				Run 2:			
Number of laboratories:		37		Number of laboratories:		37	
Arithmetic mean value:		5.760		Arithmetic mean value:		6.847	
Median:		5.844		Median:		6.791	
Standard deviation		0.789		Standard deviation		0.794	
Rel. st. deviation (%)		13.690		Rel. st. deviation (%)		11.595	
Results in decreasing order:				Results in decreasing order:			
129	27.000 (*)	146	5.844	129	26.500 (*)	171	6.791
113	8.911	36	5.790	113	10.106	8	6.790
121	6.570	171	5.727	121	7.940	15	6.790
2	6.450	1	5.710	45	7.727	1	6.670
38	6.300	109	5.650	3	7.496	36	6.610
45	6.239	169	5.630	2	7.461	115	6.610
3	6.226	15	5.580	38	7.400	23	6.600
31	6.180	6	5.530	161	7.300		
112	6.170	4	5.500	33	7.200	108	6.500
33	6.100	118	5.500	31	7.180	118	6.500
127	6.100	115	5.450	112	7.130	6	6.400
161	6.100	179	5.350	132	7.053	169	6.390
16	6.000	168	5.300	146	7.002	109	6.330
39	6.000	114	5.100	16	7.000	4	6.300
14	5.960	108	4.900	39	7.000	168	6.280
5	5.957	110	4.430	14	6.960	179	6.160
132	5.924	159	4.300	127	6.900	114	6.000
8	5.870	23	4.000	125	6.890	110	5.480
125	5.860			5	6.885	159	5.000

Table 6: Analytical results for Ni in synthetic precipitation samples, 2009.

Nickel				Nickel			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.700		Theoretical value:		0.800	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		28		Number of laboratories:		28	
Arithmetic mean value:		0.714		Arithmetic mean value:		0.747	
Median:		0.655		Median:		0.755	
Standard deviation		0.274		Standard deviation		0.114	
Rel. st. deviation (%)		38.371		Rel. st. deviation (%)		15.290	
Run 2:				Run 2:			
Number of laboratories:		27		Number of laboratories:		26	
Arithmetic mean value:		0.666		Arithmetic mean value:		0.743	
Median:		0.651		Median:		0.755	
Standard deviation		0.109		Standard deviation		0.081	
Rel. st. deviation (%)		16.393		Rel. st. deviation (%)		10.971	
Results in decreasing order:				Results in decreasing order:			
129 < 88				129 < 88			
132 < 5				132 < 5			
4 < 3.0				4 < 3.0			
108 < 2.35				108 < 2.35			
127 < 2				127 < 2			
23 2.000 (*) 5 0.651				113 < 1.5			
113 < 1.5				159 < 1.0			
159 < 1.0				168 0.920 112 0.640			
168 0.920 112 0.640				118 0.900 169 0.639			
118 0.900 169 0.639				31 0.840 110 0.630			
31 0.840 110 0.630				33 0.800 3 0.624			
33 0.800 3 0.624				146 0.773 36 0.620			
146 0.773 36 0.620				1 0.740 16 0.600			
1 0.740 16 0.600				2 0.696 114 0.600			
2 0.696 114 0.600				125 0.683 179 0.600			
125 0.683 179 0.600				115 0.682 14 0.595			
115 0.682 14 0.595				8 0.680 6 0.570			
8 0.680 6 0.570				171 0.678 15 0.570			
171 0.678 15 0.570				38 0.660 121 0.500			
38 0.660 121 0.500				161 0.660 45 0.436			
161 0.660 45 0.436				118 1.100 (*) 161 0.750			
Nickel				Nickel			
Sample no.: H3				Sample no.: H4			
Theoretical value:		9.000		Theoretical value:		8.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		38		Number of laboratories:		38	
Arithmetic mean value:		8.310		Arithmetic mean value:		7.579	
Median:		8.769		Median:		7.680	
Standard deviation		1.498		Standard deviation		0.695	
Rel. st. deviation (%)		18.031		Rel. st. deviation (%)		9.163	
Run 2:				Run 2:			
Number of laboratories:		37		Number of laboratories:		35	
Arithmetic mean value:		8.511		Arithmetic mean value:		7.734	
Median:		8.783		Median:		7.798	
Standard deviation		0.851		Standard deviation		0.460	
Rel. st. deviation (%)		9.998		Rel. st. deviation (%)		5.949	
Results in decreasing order:				Results in decreasing order:			
129 < 88				129 < 88			
132 9.592 171 8.755				45 8.773 36 7.660			
127 9.400 8 8.730				108 8.600 8 7.650			
2 9.339 36 8.630				146 8.248 115 7.650			
45 9.220 14 8.600				2 8.203 121 7.600			
112 9.210 23 8.600				127 8.200			
31 9.140 161 8.600				31 8.150 1 7.460			
33 9.100 1 8.590				112 8.120 15 7.460			
118 9.100 168 8.330				33 8.100 4 7.400			
5 9.094 6 8.290				125 8.070 6 7.380			
146 9.093 169 8.250				5 8.067 110 7.370			
16 9.000 4 8.200				16 8.000 160 7.220			
39 9.000 160 8.120				39 8.000 169 7.160			
38 8.900 109 7.650				3 7.917 132 7.141			
108 8.900 114 7.500				38 7.900 23 7.100			
15 8.810 179 6.760				118 7.900 114 6.800			
121 8.800 110 6.410				14 7.800 109 6.780			
125 8.800 113 5.926				168 7.800 179 5.960 (*)			
3 8.783 115 0.860 (*)				171 7.798 159 5.800 (*)			
				161 7.700 113 5.572 (*)			

Table 7: Analytical results for Cu in synthetic precipitation samples, 2009.

Copper				Copper			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.800		Theoretical value:		1.400	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		36		Number of laboratories:		37	
Arithmetic mean value:		1.378		Arithmetic mean value:		2.024	
Median:		0.800		Median:		1.353	
Standard deviation		1.538		Standard deviation		2.293	
Rel. st. deviation (%)		111.616		Rel. st. deviation (%)		113.287	
Run 2:				Run 2:			
Number of laboratories:		33		Number of laboratories:		36	
Arithmetic mean value:		0.960		Arithmetic mean value:		1.691	
Median:		0.800		Median:		1.351	
Standard deviation		0.600		Standard deviation		1.093	
Rel. st. deviation (%)		62.504		Rel. st. deviation (%)		64.637	
Results in decreasing order:				Results in decreasing order:			
112	<	4		112	<	4	
3	<	2.5		3	<	2.5	
132	<	1		132	<	1	
159	<	1		159	<	1	
110	7.140 (*)	33	0.800	129	14.000 (*)	112	1.350
108	5.400 (*)	127	0.800	108	5.600	160	1.340
		8	0.790			8	1.330
129	3.600	14	0.790	109	3.650	36	1.330
109	2.230	31	0.790	110	3.020	146	1.327
23	2.000	112	0.790	13	2.200	14	1.300
13	1.400	179	0.780	23	2.000	169	1.290
174	1.350	169	0.762	31	1.570	6	1.270
38	0.860	3	0.759	38	1.500	2	1.264
125	0.848	39	0.750	5	1.434	174	1.240
115	0.843	114	0.700	161	1.420	39	1.200
1	0.839	118	0.700	125	1.410	118	1.200
146	0.839	6	0.690	16	1.400	114	1.100
36	0.830	168	0.670	33	1.400	121	1.080
161	0.830	15	0.650	127	1.400	168	1.080
5	0.829	2	0.621	115	1.370	179	1.070
171	0.806	121	0.600	1	1.360	15	1.060
16	0.800	45	0.335	3	1.354	45	1.009
				171	1.353		
Copper				Copper			
Sample no.: H3				Sample no.: H4			
Theoretical value:		8.000		Theoretical value:		9.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		41		Number of laboratories:		41	
Arithmetic mean value:		7.930		Arithmetic mean value:		8.680	
Median:		7.650		Median:		8.520	
Standard deviation		1.651		Standard deviation		1.548	
Rel. st. deviation (%)		20.826		Rel. st. deviation (%)		17.834	
Run 2:				Run 2:			
Number of laboratories:		38		Number of laboratories:		39	
Arithmetic mean value:		7.529		Arithmetic mean value:		8.407	
Median:		7.577		Median:		8.510	
Standard deviation		0.801		Standard deviation		0.978	
Rel. st. deviation (%)		10.634		Rel. st. deviation (%)		11.635	
Results in decreasing order:				Results in decreasing order:			
129	14.200 (*)	36	7.580	108	14.000 (*)	36	8.510
108	12.400 (*)	146	7.574			16	8.500
		112	7.570	109	10.550	146	8.453
109	9.600	39	7.500	45	9.894	110	8.440
2	8.670	118	7.500	13	9.500	112	8.440
38	8.500	179	7.500	38	9.500	118	8.400
127	8.300	6	7.410	2	9.440	6	8.300
33	8.200	14	7.400	33	9.300	14	8.210
5	8.197	110	7.400	5	9.258	23	8.200
15	8.050	169	7.350	127	9.200	169	8.090
16	8.000	23	7.300	3	9.097	121	7.800
45	7.986	121	7.300	39	9.000	160	7.760
3	7.983	160	7.100	15	8.940	114	7.700
125	7.960	13	6.800	125	8.930	4	7.600
31	7.900	114	6.800	161	8.900	168	7.450
161	7.900	168	6.530	115	8.810	113	7.211
115	7.850	174	6.400	31	8.790	159	6.600
4	7.800	159	5.900	171	8.634	132	6.553
8	7.800	132	5.887	179	8.620	174	6.200
171	7.675	113	5.298	8	8.590	129	6.000
1	7.650			1	8.520		

Table 8: Analytical results for Zn in synthetic precipitation samples, 2009.

Zinc				Zinc			
Sample no.: H1				Sample no.: H2			
Theoretical value:	6.000			Theoretical value:	8.000		
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	37			Number of laboratories:	38		
Arithmetic mean value:	7.280			Arithmetic mean value:	11.188		
Median:	6.500			Median:	8.235		
Standard deviation	2.733			Standard deviation	18.824		
Rel. st. deviation (%)	37.537			Rel. st. deviation (%)	168.252		
Run 2:				Run 2:			
Number of laboratories:	35			Number of laboratories:	37		
Arithmetic mean value:	6.704			Arithmetic mean value:	8.139		
Median:	6.459			Median:	8.170		
Standard deviation	1.087			Standard deviation	1.056		
Rel. st. deviation (%)	16.217			Rel. st. deviation (%)	12.974		
Results in decreasing order:				Results in decreasing order:			
109 < 20				129 124.000 (*) 112 8.170			
23 20.000 (*) 171 6.459				109 < 20			
129 14.710 (*) 1 6.440				160 10.440 1 8.130			
14 10.200 161 6.400				179 9.790 169 8.060			
179 8.530 169 6.380				6 9.530 3 8.049			
160 8.360 5 6.368				108 9.400 8 8.000			
33 8.000 112 6.320				45 9.208 33 8.000			
108 7.900 146 6.274				13 9.200 5 7.991			
110 7.730 8 6.200				31 9.000 120 7.700			
6 7.380 120 6.200				125 8.600 15 7.660			
13 7.100 114 6.100				115 8.590 16 7.500			
31 7.000 121 6.100				110 8.550 168 7.400			
36 6.880 15 6.030				127 8.500 38 7.100			
125 6.880 39 6.000				36 8.460 114 7.100			
115 6.740 168 5.800				2 8.418 132 7.045			
45 6.639 38 5.300				121 8.400 23 7.000			
2 6.543 132 5.240				14 8.390 146 6.941			
16 6.500 118 4.000				171 8.369 113 5.552			
127 6.500 159 < 1				161 8.300 118 5.200			
				159 < 1			
Zinc				Zinc			
Sample no.: H3				Sample no.: H4			
Theoretical value:	100.000			Theoretical value:	110.000		
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	40			Number of laboratories:	40		
Arithmetic mean value:	99.746			Arithmetic mean value:	108.328		
Median:	99.500			Median:	109.500		
Standard deviation	10.324			Standard deviation	18.697		
Rel. st. deviation (%)	10.350			Rel. st. deviation (%)	17.260		
Run 2:				Run 2:			
Number of laboratories:	39			Number of laboratories:	37		
Arithmetic mean value:	100.919			Arithmetic mean value:	110.385		
Median:	99.500			Median:	110.000		
Standard deviation	7.273			Standard deviation	7.581		
Rel. st. deviation (%)	7.207			Rel. st. deviation (%)	6.868		
Results in decreasing order:				Results in decreasing order:			
45 119.958				160 159.000 (*) 14 109.000			
129 116.000 8 99.000				45 138.292 112 108.900			
160 114.600 31 99.000				6 121.100 13 108.000			
110 110.000 161 98.400				110 120.000 16 108.000			
6 109.500 171 97.975				125 120.000 161 107.900			
146 109.300 5 97.864				146 118.400 171 107.636			
125 109.000 14 97.800				2 116.758 120 107.600			
2 108.140 118 97.500				115 116.700 132 107.150			
127 106.000 120 97.500				121 114.000 5 107.136			
115 105.300 132 97.120				127 114.000 8 107.000			
121 104.000 16 97.000				3 113.950 15 106.000			
33 103.000 168 97.000				33 113.000 109 105.580			
3 102.900 169 97.000				108 111.500 168 104.000			
179 100.800 109 95.780				169 104.000			
1 100.000 15 94.800				1 111.000 118 103.900			
13 100.000 23 93.000				179 110.450 113 101.077			
39 100.000 113 91.202				23 110.000 38 98.700			
36 99.600 38 86.200				31 110.000 114 92.000			
112 99.600 114 85.000				36 110.000 159 65.000 (*)			
108 99.500 159 54.000 (*)				39 110.000 129 24.880 (*)			

Table 9: Analytical results for As in synthetic precipitation samples, 2009.

Arsenic				Arsenic			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.700		Theoretical value:		0.500	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		26		Number of laboratories:		26	
Arithmetic mean value:		0.682		Arithmetic mean value:		0.532	
Median:		0.680		Median:		0.485	
Standard deviation		0.148		Standard deviation		0.209	
Rel. st. deviation (%)		21.661		Rel. st. deviation (%)		39.386	
Run 2:				Run 2:			
Number of laboratories:		24		Number of laboratories:		25	
Arithmetic mean value:		0.677		Arithmetic mean value:		0.493	
Median:		0.680		Median:		0.480	
Standard deviation		0.076		Standard deviation		0.071	
Rel. st. deviation (%)		11.158		Rel. st. deviation (%)		14.385	
Results in decreasing order:				Results in decreasing order:			
113 < 4.000				113 < 4.000			
4 < 1.500				4 < 1.500			
118	1.200 (*)	8	0.680	118	1.500 (*)	8	0.480
159 < 1.000				159 < 1.000			
115	0.892	31	0.680	115	0.735	31	0.480
179	0.770	3	0.671	23	0.600	1	0.476
125	0.769	38	0.670	121	0.560	146	0.461
6	0.730	15	0.640	125	0.544	36	0.460
121	0.730	5	0.633	6	0.520	3	0.458
2	0.705	169	0.626	33	0.520	15	0.450
171	0.705	168	0.620	161	0.510	110	0.450
127	0.700	14	0.602	14	0.506	168	0.450
33	0.690	146	0.572	2	0.504	179	0.450
36	0.690	45	0.555	127	0.500	169	0.446
161	0.690	110	0.540	171	0.500	5	0.434
1	0.680	23	0.300 (*)	38	0.490	45	0.337
Arsenic				Arsenic			
Sample no.: H3				Sample no.: H4			
Theoretical value:		6.000		Theoretical value:		7.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		31		Number of laboratories:		31	
Arithmetic mean value:		5.869		Arithmetic mean value:		6.901	
Median:		5.840		Median:		6.895	
Standard deviation		0.393		Standard deviation		0.561	
Rel. st. deviation (%)		6.698		Rel. st. deviation (%)		8.133	
Run 2:				Run 2:			
Number of laboratories:		29		Number of laboratories:		28	
Arithmetic mean value:		5.932		Arithmetic mean value:		6.855	
Median:		5.900		Median:		6.857	
Standard deviation		0.318		Standard deviation		0.419	
Rel. st. deviation (%)		5.357		Rel. st. deviation (%)		6.111	
Results in decreasing order:				Results in decreasing order:			
115	6.530	38	5.800	45	8.279 (*)	3	6.820
125	6.500	113	5.758	113	8.025 (*)	38	6.800
45	6.481	171	5.748	146	7.808	31	6.750
159	6.300	3	5.740	125	7.700	14	6.710
6	6.250	121	5.690	161	7.400	171	6.691
2	6.240	169	5.690	115	7.320	8	6.680
146	6.224	179	5.650	159	7.200	118	6.500
118	6.200	31	5.630	6	7.170	1	6.490
127	6.200	1	5.590	15	7.150	169	6.470
161	6.040	168	5.560	127	7.100	5	6.469
15	6.000	5	5.513	33	7.000	121	6.380
33	6.000	23	5.500	39	7.000	179	6.360
39	6.000	14	5.460	36	6.990	168	6.290
36	5.990	114	5.000 (*)	4	6.900	114	6.000
4	5.900	110	4.900 (*)	23	6.900	110	5.680 (*)
8	5.840			2	6.895		

Table 10: Analytical results for Cd in synthetic precipitation samples, 2009.

Cadmium				Cadmium			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.070		Theoretical value:		0.060	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		31		Number of laboratories:		30	
Arithmetic mean value:		0.075		Arithmetic mean value:		0.056	
Median:		0.070		Median:		0.059	
Standard deviation		0.045		Standard deviation		0.017	
Rel. st. deviation (%)		59.525		Rel. st. deviation (%)		29.705	
Run 2:				Run 2:			
Number of laboratories:		30		Number of laboratories:		27	
Arithmetic mean value:		0.067		Arithmetic mean value:		0.058	
Median:		0.070		Median:		0.060	
Standard deviation		0.016		Standard deviation		0.007	
Rel. st. deviation (%)		23.335		Rel. st. deviation (%)		12.304	
Results in decreasing order:				Results in decreasing order:			
132	<	5		132	<	5	
129	<	4		129	<	4	
159	<	1		159	<	1	
23	<	0.8		23	<	0.8	
113	<	0.4		113	<	0.4	
4	<	0.35		4	<	0.35	
127	<	0.2		127	<	0.2	
23	0.300	(*)		114	0.100	(*)	
1	<	0.1		1	<	0.1	
6	<	0.1		6	<	0.1	
109	<	0.1	179 0.070	109	<	0.1	15 0.059
114	0.100	14 0.069		13	0.070	31 0.059	
174	0.094	2 0.067		14	0.068	38 0.058	
121	0.090	5 0.067		112	0.067	115 0.058	
146	0.080	3 0.066		174	0.067	171 0.057	
112	0.077	169 0.064		125	0.062	5 0.056	
8	0.073	36 0.063		8	0.061	3 0.054	
115	0.073	141 0.062		146	0.061	169 0.052	
125	0.072	13 0.060		2	0.060	168 0.050	
31	0.071	38 0.060		16	0.060	179 0.050	
15	0.070	110 0.060		33	0.060	36 0.045	
16	0.070	168 0.060		39	0.060	141 0.045	
33	0.070	10 0.059		110	0.060	10 0.040	
39	0.070	45 0.024		121	0.060	45 0.017	(*)
161	0.070	118 0.020		161	0.060	118 0.000	(*)
171	0.070						
Cadmium				Cadmium			
Sample no.: H3				Sample no.: H4			
Theoretical value:		0.700		Theoretical value:		0.600	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		37		Number of laboratories:		37	
Arithmetic mean value:		0.665		Arithmetic mean value:		0.577	
Median:		0.673		Median:		0.590	
Standard deviation		0.064		Standard deviation		0.081	
Rel. st. deviation (%)		9.642		Rel. st. deviation (%)		13.998	
Run 2:				Run 2:			
Number of laboratories:		34		Number of laboratories:		36	
Arithmetic mean value:		0.671		Arithmetic mean value:		0.587	
Median:		0.676		Median:		0.591	
Standard deviation		0.048		Standard deviation		0.050	
Rel. st. deviation (%)		7.226		Rel. st. deviation (%)		8.574	
Results in decreasing order:				Results in decreasing order:			
132	<	5		132	<	5	
129	<	4		129	<	4	
159	<	1		159	<	1	
127	0.800	(*)	1 0.672	13	0.710	36 0.590	
113	0.774	31 0.670		110	0.680	171 0.588	
125	0.723	118 0.670		113	0.675	10 0.587	
2	0.721	5 0.669		125	0.631	8 0.581	
121	0.720	14 0.669		121	0.630	6 0.580	
146	0.719	13 0.660		161	0.620	5 0.579	
115	0.713	15 0.660		146	0.619	14 0.579	
112	0.710	169 0.643		115	0.618	1 0.576	
16	0.700	38 0.640		112	0.613	38 0.570	
33	0.700	45 0.636		45	0.612	31 0.560	
36	0.700	4 0.610		15	0.600	118 0.560	
39	0.700	23 0.600		16	0.600	169 0.541	
161	0.700	114 0.600		23	0.600	141 0.501	
168	0.700	141 0.594		39	0.600	4 0.500	
6	0.690	179 0.570		127	0.600	114 0.500	
10	0.690	110 0.550		168	0.600	179 0.500	
171	0.681	174 0.510	(*)	3	0.594	174 0.470	
8	0.679	109 0.500	(*)	2	0.593	109 0.200	(*)
3	0.673			33	0.590		

Table 11: Analytical results for Pb in synthetic precipitation samples, 2009.

Lead		Lead	
Sample no.: H1		Sample no.: H2	
Theoretical value:	1.500	Theoretical value:	1.000
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	34	Number of laboratories:	33
Arithmetic mean value:	1.630	Arithmetic mean value:	0.967
Median:	1.495	Median:	1.000
Standard deviation	1.164	Standard deviation	0.287
Rel. st. deviation (%)	71.395	Rel. st. deviation (%)	29.660
Run 2:		Run 2:	
Number of laboratories:	33	Number of laboratories:	30
Arithmetic mean value:	1.437	Arithmetic mean value:	0.990
Median:	1.490	Median:	1.000
Standard deviation	0.301	Standard deviation	0.107
Rel. st. deviation (%)	20.911	Rel. st. deviation (%)	10.801
Results in decreasing order:		Results in decreasing order:	
129 8.000 (*) 31 1.490		110 1.910 (*) 127 1.000	
113 < 7		113 < 7	
132 < 5		132 < 5	
10 2.027 168 1.480		39 1.300 112 0.990	
23 2.000 121 1.460		179 1.150 36 0.970	
13 1.800 8 1.450		13 1.100 8 0.960	
110 1.650 3 1.429		14 1.070 2 0.934	
179 1.650 112 1.420		125 1.060 171 0.934	
115 1.540 14 1.410		121 1.050 3 0.928	
125 1.540 171 1.403		168 1.050 169 0.923	
161 1.540 15 1.400		10 1.047 38 0.920	
146 1.534 169 1.390		161 1.030 23 0.900	
36 1.530 2 1.356		5 1.029 33 0.900	
5 1.524 38 1.300		31 1.020 114 0.900	
1 1.500 114 1.300		115 1.020 45 0.763	
16 1.500 6 1.130		146 1.018 6 0.720	
33 1.500 45 1.120		1 1.010 174 0.290 (*)	
39 1.500 4 < 1		4 < 1 118 0.000 (*)	
127 1.500 159 < 1		159 < 1	
118 0.600		15 1.000	
174 0.450		16 1.000	
Lead		Lead	
Sample no.: H3		Sample no.: H4	
Theoretical value:	25.000	Theoretical value:	30.000
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	40	Number of laboratories:	40
Arithmetic mean value:	23.290	Arithmetic mean value:	28.942
Median:	24.301	Median:	29.930
Standard deviation	3.631	Standard deviation	3.706
Rel. st. deviation (%)	15.591	Rel. st. deviation (%)	12.807
Run 2:		Run 2:	
Number of laboratories:	38	Number of laboratories:	38
Arithmetic mean value:	24.042	Arithmetic mean value:	29.228
Median:	24.385	Median:	29.930
Standard deviation	1.402	Standard deviation	1.949
Rel. st. deviation (%)	5.833	Rel. st. deviation (%)	6.670
Results in decreasing order:		Results in decreasing order:	
127 26.400 2 24.232		129 36.400 (*) 146 29.860	
132 26.340 13 24.100		110 33.300 112 29.210	
160 25.730 4 24.000		132 32.640 31 29.160	
6 25.720 8 23.600		179 31.620 113 28.735	
115 25.600 36 23.400		160 31.540 23 28.500	
121 25.600 168 23.400		6 31.040 4 28.400	
5 25.279 171 23.260		127 30.900 2 28.380	
16 25.000 169 23.100		115 30.700 38 28.300	
33 25.000 23 23.000		161 30.700 8 28.000	
1 24.900 39 23.000		5 30.522 36 28.000	
14 24.900 159 23.000		121 30.400 39 28.000	
15 24.900 113 22.792		3 30.395 171 27.753	
3 24.890 38 22.500		118 30.300 45 27.570	
31 24.880 110 22.300		1 30.200 168 27.200	
118 24.800 45 21.808		13 30.200 169 27.100	
125 24.700 109 21.750		14 30.000 159 27.000	
146 24.620 10 21.380		15 30.000 109 25.650	
112 24.460 114 20.500		16 30.000 114 25.000	
161 24.400 174 11.400 (*)		33 30.000 10 24.400	
179 24.370 129 6.600 (*)		125 30.000 174 10.600 (*)	




Figure 1: Youden plot of chromium, 2009.




Figure.2: Youden plot of nickel, 2009.




Figure.3: Youden plot of copper, 2009.




Figure.4: Youden plot of zinc, 2009.




Figure.5: Youden plot of arsenic, 2009.




Figure 6: Youden plot of cadmium, 2009.




Figure 7: Youden plot of lead, 2009.

Appendix 2

Tables and figures, 2010

Table 12a: Participating laboratories in the EMEP network, 2010. The numbers in front are used in tables.

No	Laboratory identification
2	Flemish Environment Agency, Belgium
3	Czech Hydrometeorological Institute, Czech Republic
4	National Environmental Research Institute. Air Pollution Laboratory, Denmark
5	Finnish Meteorological Institute, Finland
8	Umweltbundesamt, Germany
10	Hungarian Meteorological Service, Hungary
13	C.N.R. Istituto Inquinamento Atmosferico, Italy
14	RIVM Laboratory of Inorganic Analytical Chemistry, The Netherlands
15	The Norwegian Institute for Air Research, Norway
16	Inst. Of Meteorology and Water Management, Poland
23	AEA Technology, National Environmental Techn. Centre, United Kingdom
24	Hydrometeorological Institute of Serbia, Serbia
31	Slovak Hydrometeorological Institute, Slovakia
32	Atmospheric Pollution Research Laboratory, Institute of Physics, Lithuania
33	Environmental Pollution Observ. Centre, Latvia
36	Hydrometeorological Institute of Slovenia, Slovenia
38	Estonian Environmental Research Centre, Estonia

Table 12b: Participating laboratories outside the EMEP network, 2010. The number in front of the names is used in tables and figures.

No	Laboratory identification
110	Thüringer Landesanstalt für Landwirtschaft (TTL), Germany
112	Niedersächsische Forstliche Versuchsanstalt (NVF), Germany
115	Bayerische Landesanstalt f. Wald- und Forstwirtschaft, Germany
118	Forstliche Versuchs- und Forschungsanstalt Baden-Württemberg, Germany
120	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA), Germany
121	Landesamt für Natur und Umwelt, Germany
125	Bayerisches Landesamt für Umweltschutz, Germany
127	Dept. Of Chemistry, Environmental Health Division, Malaysia
129	Ecole Nationale d'ingénieurs de Sfax, Tunisia
132	Comision Chilena De Energia Nuclear, Chile
141	Pollutants Chemical Analysis Centre, Marine Division, Japan
159	CARSO, France
161	National Institute of Chemistry, Slovenia
168	Universite de Bretagne Occidentale, France
169	Lancaster Environment Centre, Centre for Ecology & Hydrology, UK
171	Ecole de Mines de Douai, Departement Chimie et Environment, France
174	Background Monitoring Department, Institute of Global Climate and Ecology(IGCE), Russian Federation
176	Main Geophysical Observatory, Russian Federation
178	Environmental Chemistry and EANET Monitoring Laboratory, Limnological Institute RAS/SB, Russian Federation
179	Murmansk Environmental Monitoring Center, Russian Federation

Table 13: Analytical techniques used at the participating laboratories for the different elements, 2010.

Lab no	Network	As	Cd	Cr	Cu	Pb	Ni	Zn
2	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
3	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
4	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
5	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
8	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
10	EMEP		GF-AAS			GF-AAS		
13	EMEP		GF-AAS		GF-AAS	GF-AAS		GF-AAS
14	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
15	EMEP	HR-ICP-MS	HR-ICP-MS	HR-ICP-MS	HR-ICP-MS	HR-ICP-MS	HR-ICP-MS	HR-ICP-MS
16	EMEP		GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS Higher conc F-AAS
23	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-AES
24	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
31	EMEP	HG-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS
32	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
33	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
36	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
38	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
110	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
112	ICP-Forest		ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
115	ICP-Forest	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
118	EMEP	ICP-AES	GF-AAS	ICP-AES	ICP-AES	GF-AAS	ICP-AES	ICP-AES
120	ICP-Forest	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	ICP-AES
121	EMEP	Hydrid-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	Voltammetry
125	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
127	EANET	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
129	EMEP	AAS	AAS	AAS	AAS	AAS	AAS	AAS
132	EMEP		ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES	ICP-AES
141	WMO-GAW		GF-AAS					
159	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
161	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
168	Other	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
169	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
171	EMEP	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
174	Other		GF-AAS	Cr methods	GF-AAS	GF-AAS	GF-AAS	GF-AAS
176	EMEP		GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS
178	EANET	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
179	EMEP	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	GF-AAS	F-AAS

Table 14: Reported results for metal determination in precipitation samples, expressed as % deviation from expected value, 2010.

Lab no	Arsenic				Cadmium				Chromium				Copper				Lead				Nickel				Zinc				
	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	H1	H2	H3	H4	
2	6	7	7	9	17	5	3	6	1	3	4	5	9	6	9	11	0	0	-1	-3	-4	1	3	-10	12	6	6	10	
3	4	0	1	0	57	7	8	12	3	3	4	1	11	6	6	9	-5	-6	-6	-3	1	3	4	-1	1	5	5	3	
4	< 1.5	-5	-4	< 1.5	< 0.35	-14	-20	< 0.35	< 2.0	3	4	< 2.0	< 2.5	-14	-13	< 2.5	< 3.0	-3	-1	< 3.0		2	2	< 3.0					
5	-10	-7	-7	-10	-8	-7	-8	-16	-9	-7	-7	-10	-8	-7	-7	-9	-7	-7	-7	-8	-3	-6	-6	-4	-6	-7	-6	-4	
8	-5	-5	-5	-5	-2	-6	-5	-6	-2	-7	-7	-2	0	-4	-4	-5	-2	-8	-9	-3	-2	-6	-6	-3	14	-1	-2	2	
10					-25	4	4	-34									-1	12	14	17									
13					17	-31	-14	20					6	-9	-17	-7	0	-7	-11	-36				58	8	-3	50		
14	-4	-2	-1	-4	8	0	5	-20	0	2	0	-9	-1	-3	-2	-5	4	-1	-1	0	-5	-1	-4	-8	9	0	-1	5	
15	-3	4	1	8	0	-2	-3	0	-1	0	1	1	-3	0	0	2	-3	-2	-1	10	1	-1	1	6	-4	-4	0	0	
16					0	0	0	0	0	0	0	0	-11	0	7	0	8	0	0	0	11	0	0	7	0	0	0	14	
23	4	-1	-3	17	13	-5	-8	29	7	9	0	65	< 2	-14	-20	< 2	41	-1	-2	81		-3	-3	< 0.9	-24	5	5	-43	
24	33	2	-2	50	100	-1	-4	380	0	-7	-10	0	3311	281	330	1040	154	-4	-8	256	157	16	6	286	514	22	24	437	
31	-8	-4	-3	-10	0	-6	-6	0	3	-1	0	4	2	-2	-1	0	-7	-8	-8	-9	1	-2	-1	-3	20	0	-3	14	
32	33	-3	-12	-25	-33	13	0	20	14	3	2	0	33	18	19	20	23	8	9	11	1	2	3	0	12	-2	-2	-4	
33	33	12	14	100	-33	-3	-16	-20	29	17	22	120	33	6	1	20	-42	4	9	-56	11	-2	-70	29	100	0	3	43	
36	8	3	4	12	33	2	5	30	0	-2	0	2	-9	-5	-2	-6	-1	-5	-6	5	-7	-3	-1	0	6	3	0	19	
38	16	-90	-1	23	3	-1	2	6	-9	-6	-4	< 0.5	< 1	2	3	2	0	-3	-1	4	3	-3	0	8	-7	3	2	-1	
110	0	0	1	0	< 0.1	-6	-4	< 0.1	-14	-9	-10	-20	11	0	0	7	21	2	3	11	18	0	-1	19	23	1	1	0	
112					0	2	2	0	-1	0	0	-8	-9	-8	-8	-9	2	-1	-1	-1	-32	0	1	-46	13	2	3	9	
115	28	1	4	38	7	3	5	6	-11	-5	-4	-11	-29	-7	-6	-24	62	8	8	12	-1	-5	-4	-3	27	5	6	8	
118		-8	-12		-33	-3	-2	-40	-29	-8	-10	-80	-11	-4	-1	-7	15	7	9	0	-9	-6	-5	-57	10	-4	-4	-11	
120	-33	-13	-10	-75	0	0	0	-20	-29	-7	-6	-40	-11	-38	-9	-13	54	-20	-17	122		0	40		22	0	0	9	
121	7	-1	-1	5	0	11	4	-20	9	1	3	10	-20	-9	-7	-5	5	0	-2	3	9	-4	0	9	8	-1	4	2	
125	7	9	13	13	2	3	7	6	-1	1	5	3	13	6	10	11	2	0	3	3	6	3	7	7	27	14	19	18	
127	-17	-3	-4	0	< 0.2	-25	-20	< 0.2	0	-5	-4	0	-11	-8	-9	-13	-8	-6	-7	-11	-6	-5	-5	-40	-6	-7	-20		
129					2133	113	< 5	< 5	< 50	< 50	< 50	< 50	1256	113	86	407	-55	24	-19	-40	443	11	0	1343	< 8	-82	< 8	< 8	
132					< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	< 10	0	9	< 10		< 10	10	< 10	< 10	< 10	< 10	< 10	< 10
141					-20	-8	-10	-24																					
159	30	5	6	55	< 0.1	< 0.1	-2	1480	-26	-8	-8	< 0.5	< 1	-5	-4	-13	-15	-4	-4	< 1		-12	-11	< 0.5	-10	10	11	-11	
161	-8	0	-12	-18	0	-8	-10	0	-11	-4	-3	-14	-6	-3	-2	-5	41	-2	-3	-6	-14	-5	-4	-17	6	0	-1	3	
168	-7	-1	-3	7	-8	3	6	2	-7	0	3	0	54	8	11	13	10	2	2	1	-12	-4	1	-10	6	-13	-7	1	
169	-3	-2	-2	-2	-3	-2	-5	-3	-4	-4	-3	0	-2	-3	-4	-3	1	-2	-2	-1	-1	-4	-3	-1	11	4	5	9	
171	7	4	4	5	13	9	10	6	-4	-3	-6	-10	6	4	4	6	6	-1	0	2	0	-1	-1	-1	31	15	15	30	
174					25	-11	-4	36					-13	-6	-4	-33	38	8	4	67		10	-12		21		33		
176					-25	-36	-36	-32	-64	-68	-69	-70	219	-45	-49	295	-70	-79	-62	-51	619	927	84	2869	503	-21	-70	268	
178	38	14	20	34	122	13	26	84	8	4	9	60	-12	11	15	-3	1	-4	-2	41	23	12	10	32	34	21	23	32	
179	28	6	13	13	17	-38	14	0	6	3	7	6	-13	8	7	-29	27	26	6	28	-19	-34	-32	7	71	10	6	40	

between ± 25–50%
more than ± 50%
for low theoretical values ((Pb, Ni, Cr, As <1 µg/l, Cd <0.5 µg/l, Cu <2 µg/l)

between ± 15–30%
more than ± 30%
for high theoretical values ((Pb, Ni, Cr, As >1 µg/l, Cd >0.5 µg/l, Cu >2 µg/l)

Table 15: Analytical results for Cr in synthetic precipitation samples, 2010.

Chromium				Chromium			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.700		Theoretical value:		6.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		30		Number of laboratories:		31	
Arithmetic mean value:		0.666		Arithmetic mean value:		5.805	
Median:		0.691		Median:		5.960	
Standard deviation		0.114		Standard deviation		0.796	
Rel. st. deviation (%)		17.112		Rel. st. deviation (%)		13.718	
Run 2:				Run 2:			
Number of laboratories:		28		Number of laboratories:		30	
Arithmetic mean value:		0.673		Arithmetic mean value:		5.935	
Median:		0.691		Median:		5.980	
Standard deviation		0.074		Standard deviation		0.340	
Rel. st. deviation (%)		10.999		Rel. st. deviation (%)		5.725	
Results in decreasing order:				Results in decreasing order:			
33	0.900 (*)	112	0.690	33	7.000	36	5.890
32	0.800	125	0.690	23	6.530	171	5.804
121	0.760	8	0.685	178	6.216	169	5.790
178	0.754	171	0.671	2	6.201	161	5.750
23	0.747	169	0.669	4	6.200	127	5.700
179	0.740	168	0.650	32	6.200	115	5.680
3	0.723	5	0.640	3	6.190	38	5.640
31	0.720	38	0.635	179	6.160	8	5.600
2	0.710	115	0.624	14	6.148	24	5.600
36	0.702	161	0.620	125	6.060	120	5.600
16	0.700	110	0.600	121	6.040	5	5.570
24	0.700	159	0.520	112	6.020	118	5.500
127	0.700	118	0.500	168	6.020	159	5.500
14	0.697	120	0.500	15	6.015	110	5.460
15	0.692	176	0.250 (*)	16	6.000	176	1.910 (*)
		129	< 50.000	31	5.960	129	< 50.000
		132	< 10.000			132	< 10.000
		4	< 2.000				
Chromium				Chromium			
Sample no.: H3				Sample no.: H4			
Theoretical value:		5.000		Theoretical value:		0.500	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		31		Number of laboratories:		28	
Arithmetic mean value:		4.860		Arithmetic mean value:		0.501	
Median:		4.980		Median:		0.500	
Standard deviation		0.694		Standard deviation		0.184	
Rel. st. deviation (%)		14.287		Rel. st. deviation (%)		36.683	
Run 2:				Run 2:			
Number of laboratories:		30		Number of laboratories:		26	
Arithmetic mean value:		4.970		Arithmetic mean value:		0.493	
Median:		4.985		Median:		0.500	
Standard deviation		0.333		Standard deviation		0.125	
Rel. st. deviation (%)		6.694		Rel. st. deviation (%)		25.329	
Results in decreasing order:				Results in decreasing order:			
33	6.100	36	4.980	33	1.100 (*)	32	0.500
178	5.446	169	4.860	23	0.827	127	0.500
179	5.350	161	4.840	178	0.801	168	0.500
125	5.240	38	4.810	121	0.550	8	0.492
4	5.200	115	4.810	179	0.530	112	0.460
2	5.192	127	4.800	2	0.523	14	0.456
3	5.190	171	4.720	31	0.520	5	0.450
121	5.170	120	4.700	125	0.516	171	0.450
168	5.130	5	4.670	36	0.512	115	0.445
32	5.100	8	4.670	3	0.506	161	0.430
15	5.062	159	4.600	15	0.504	110	0.400
112	5.020	24	4.500	169	0.501	120	0.300
14	5.001	118	4.500	16	0.500	176	0.150
16	5.000	110	4.480	24	0.500	118	0.100 (*)
31	4.990	176	1.560 (*)			129	< 50.000
23	4.980	129	< 50.000			132	< 10.000
		132	< 10.000			4	< 2.000
						38	< 0.500
						159	< 0.500

Table 16: Analytical results for Ni in synthetic precipitation samples, 2010.

Nickel				Nickel			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.791		Theoretical value:		9.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		27		Number of laboratories:		33	
Arithmetic mean value:		1.244		Arithmetic mean value:		11.397	
Median:		0.800		Median:		8.870	
Standard deviation		1.493		Standard deviation		14.567	
Rel. st. deviation (%)		120.033		Rel. st. deviation (%)		127.814	
Run 2:				Run 2:			
Number of laboratories:		25		Number of laboratories:		32	
Arithmetic mean value:		0.846		Arithmetic mean value:		8.865	
Median:		0.791		Median:		8.835	
Standard deviation		0.344		Standard deviation		0.748	
Rel. st. deviation (%)		40.644		Rel. st. deviation (%)		8.442	
Results in decreasing order:				Results in decreasing order:			
176	7.120 (*)	171	0.791	176	92.440 (*)	31	8.800
129	5.320 (*)	169	0.785	24	10.400	33	8.800
24	2.400	115	0.782	178	10.122	23	8.750
178	1.022	8	0.773	129	10.000	38	8.710
110	0.970	5	0.760	174	9.900	36	8.700
16	0.900	2	0.752	3	9.300	121	8.670
33	0.900	14	0.744	125	9.250	169	8.660
121	0.880	36	0.723	4	9.200	168	8.630
125	0.850	118	0.700	32	9.200	161	8.580
38	0.819	168	0.670	2	9.072	115	8.540
15	0.805	161	0.650	112	9.020	118	8.500
3	0.804	179	0.600	16	9.000	127	8.500
31	0.800	112	0.460	120	9.000	5	8.470
32	0.800	132	< 10.000	110	8.980	8	8.420
		4	< 3.000	14	8.892	159	7.900
		127	< 2.000	171	8.872	179	5.960
		174	< 1.000	15	8.870	132	< 10
		23	< 0.900				
		159	< 0.500				
Nickel				Nickel			
Sample no.: 3				Sample no.: 4			
Theoretical value:		10.000		Theoretical value:		0.700	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		34		Number of laboratories:		27	
Arithmetic mean value:		10.009		Arithmetic mean value:		1.855	
Median:		9.967		Median:		0.700	
Standard deviation		2.158		Standard deviation		4.206	
Rel. st. deviation (%)		21.562		Rel. st. deviation (%)		226.716	
Run 2:				Run 2:			
Number of laboratories:		32		Number of laboratories:		26	
Arithmetic mean value:		9.966		Arithmetic mean value:		1.127	
Median:		9.967		Median:		0.697	
Standard deviation		1.051		Standard deviation		1.876	
Rel. st. deviation (%)		10.549		Rel. st. deviation (%)		166.455	
Results in decreasing order:				Results in decreasing order:			
176	18.390 (*)	171	9.944	176	20.780 (*)	171	0.694
120	14.000	110	9.910	129	10.100	169	0.692
178	11.040	36	9.890	24	2.700	3	0.690
132	11.000	31	9.880	178	0.923	115	0.682
125	10.700	169	9.730	33	0.900	8	0.681
24	10.600	23	9.720	110	0.830	31	0.680
3	10.400	115	9.640	121	0.760	5	0.670
32	10.300	161	9.610	38	0.757	14	0.642
2	10.260	14	9.598	16	0.750	168	0.630
4	10.200	118	9.500	179	0.750	2	0.629
15	10.100	127	9.500	125	0.746	161	0.580
112	10.070	8	9.430	15	0.742	112	0.380
168	10.050	5	9.380	36	0.702	118	0.300
16	10.000	159	8.900	32	0.700	132	< 10.000
121	10.000	174	8.800			4	< 3.000
129	10.000	179	6.760			127	< 2.000
38	9.990	33	3.000 (*)			174	< 1.000
						23	< 0.900
						159	< 0.500

Table 17: Analytical results for Cu in synthetic precipitation samples, 2010.

Copper				Copper			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.900		Theoretical value:		8.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		30		Number of laboratories:		34	
Arithmetic mean value:		2.338		Arithmetic mean value:		8.651	
Median:		0.885		Median:		7.768	
Standard deviation		5.748		Standard deviation		4.282	
Rel. st. deviation (%)		245.829		Rel. st. deviation (%)		49.496	
Run 2:				Run 2:			
Number of laboratories:		29		Number of laboratories:		33	
Arithmetic mean value:		1.360		Arithmetic mean value:		7.989	
Median:		0.878		Median:		7.747	
Standard deviation		2.122		Standard deviation		1.881	
Rel. st. deviation (%)		156.005		Rel. st. deviation (%)		23.543	
Results in decreasing order:				Results in decreasing order:			
24	30.700 (*)	169	0.878	24	30.500 (*)	14	7.747
129	12.200	15	0.871	129	17.000	169	7.730
176	2.870	161	0.850	32	9.400	118	7.700
168	1.390	5	0.830	178	8.842	8	7.680
32	1.200	112	0.820	168	8.660	36	7.630
33	1.200	36	0.815	179	8.620	159	7.600
125	1.020	16	0.800	33	8.500	174	7.500
3	1.000	118	0.800	125	8.500	115	7.470
110	1.000	120	0.800	2	8.483	5	7.420
2	0.983	127	0.800	3	8.480	127	7.400
171	0.955	178	0.789	171	8.335	112	7.390
13	0.950	174	0.780	38	8.170	13	7.300
31	0.920	179	0.780	15	8.010	121	7.260
8	0.900	121	0.720	16	8.000	4	6.900
14	0.892	115	0.636	110	7.980	23	6.880
		132	< 10.000	31	7.840	120	5.000
		4	< 2.500	161	7.790	176	4.420
		23	< 2.000			132	< 10.000
		38	< 1.000				
		159	< 1.000				
Copper				Copper			
Sample no.: H3				Sample no.: H4			
Theoretical value:		7.000		Theoretical value:		1.500	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		34		Number of laboratories:		32	
Arithmetic mean value:		7.697		Arithmetic mean value:		2.274	
Median:		6.888		Median:		1.458	
Standard deviation		4.179		Standard deviation		3.017	
Rel. st. deviation (%)		54.291		Rel. st. deviation (%)		132.635	
Run 2:				Run 2:			
Number of laboratories:		33		Number of laboratories:		31	
Arithmetic mean value:		7.018		Arithmetic mean value:		1.796	
Median:		6.875		Median:		1.455	
Standard deviation		1.360		Standard deviation		1.356	
Rel. st. deviation (%)		19.376		Rel. st. deviation (%)		75.521	
Results in decreasing order:				Results in decreasing order:			
24	30.100 (*)	14	6.875	24	17.100 (*)	178	1.455
129	13.000	36	6.840	129	7.600	8	1.430
32	8.300	161	6.830	176	5.930	161	1.430
178	8.017	8	6.700	32	1.800	14	1.426
168	7.780	159	6.700	33	1.800	121	1.420
125	7.730	169	6.700	168	1.690	36	1.410
2	7.611	174	6.700	125	1.660	13	1.400
16	7.500	115	6.610	2	1.659	118	1.400
179	7.500	5	6.520	3	1.630	5	1.370
3	7.450	121	6.480	110	1.600	112	1.360
171	7.268	112	6.440	171	1.585	120	1.300
38	7.230	120	6.400	38	1.530	127	1.300
33	7.100	127	6.400	15	1.523	159	1.300
15	7.031	4	6.100	16	1.500	115	1.140
110	6.980	13	5.800	31	1.500	179	1.070
31	6.940	23	5.600	169	1.460	174	1.000
118	6.900	176	3.570			132	< 10.000
		132	< 10.000			4	< 2.500
						23	< 2.000

Table 18: Analytical results for Zn in synthetic precipitation samples, 2010.

Zinc		Zinc	
Sample no.: H1		Sample no.: H2	
Theoretical value:	5.000	Theoretical value:	100.000
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	31	Number of laboratories:	34
Arithmetic mean value:	7.332	Arithmetic mean value:	100.219
Median:	5.575	Median:	100.550
Standard deviation	6.299	Standard deviation	17.186
Rel. st. deviation (%)	85.920	Rel. st. deviation (%)	17.149
Run 2:		Run 2:	
Number of laboratories:	29	Number of laboratories:	33
Arithmetic mean value:	5.740	Arithmetic mean value:	102.717
Median:	5.560	Median:	101.000
Standard deviation	1.351	Standard deviation	9.267
Rel. st. deviation (%)	23.539	Rel. st. deviation (%)	9.022
Results in decreasing order:		Results in decreasing order:	
24 30.700 (*) 169 5.560		24 122.300 14 100.100	
176 30.130 (*) 118 5.500		174 121.000 16 100.000	
33 10.000 14 5.450		178 120.690 31 100.000	
179 8.530 121 5.420		171 115.400 33 100.000	
13 7.900 36 5.320		125 114.000 120 99.800	
178 6.675 168 5.300		179 110.450 161 99.710	
171 6.536 161 5.280		159 110.000 8 98.900	
115 6.370 3 5.030		13 108.000 121 98.700	
125 6.370 16 5.000		2 106.400 32 98.000	
110 6.150 15 4.795		3 105.000 15 96.350	
120 6.100 5 4.700		23 105.000 118 95.600	
31 6.000 38 4.660		115 104.900 127 93.800	
8 5.680 159 4.500		169 104.000 5 93.400	
112 5.630 23 3.820		36 103.000 132 92.700	
32 5.600 127 3.000		38 103.000 168 87.500	
2 5.575 132 < 10.000		112 101.800 176 79.150	
		110 101.000 129 17.800 (*)	
Zinc		Zinc	
Sample no.: H3		Sample no.: H4	
Theoretical value:	95.000	Theoretical value:	7.000
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	33	Number of laboratories:	31
Arithmetic mean value:	96.735	Arithmetic mean value:	9.103
Median:	96.200	Median:	7.580
Standard deviation	15.083	Standard deviation	6.351
Rel. st. deviation (%)	15.592	Rel. st. deviation (%)	69.769
Run 2:		Run 2:	
Number of laboratories:	32	Number of laboratories:	29
Arithmetic mean value:	98.855	Arithmetic mean value:	7.545
Median:	96.600	Median:	7.383
Standard deviation	9.042	Standard deviation	1.330
Rel. st. deviation (%)	9.147	Rel. st. deviation (%)	17.633
Results in decreasing order:		Results in decreasing order:	
174 126.000 120 95.400		24 37.600 (*) 14 7.383	
24 117.700 36 95.200		176 25.780 (*) 3 7.200	
178 117.260 16 95.000		13 10.500 161 7.190	
125 113.000 15 94.630		33 10.000 121 7.150	
171 109.210 161 94.440		179 9.790 8 7.120	
159 105.000 14 94.130		178 9.229 168 7.100	
179 100.800 8 93.300		171 9.066 15 7.022	
2 100.700 32 93.000		36 8.350 110 7.000	
115 100.700 13 92.000		125 8.270 38 6.940	
3 100.000 31 92.000		16 8.000 5 6.700	
169 99.900 118 91.600		31 8.000 32 6.700	
23 99.700 5 89.200		2 7.668 118 6.200	
121 99.200 168 88.800		112 7.630 159 6.200	
33 98.000 127 88.300		169 7.610 127 5.600	
112 97.900 132 88.100		120 7.600 23 4.020	
38 97.000 176 28.900 (*)		115 7.580 132 < 10.000	
110 96.200 129 < 8.000			

Table 19: Analytical results for As in synthetic precipitation samples, 2010.

Arsenic				Arsenic			
Sample no.: H1				Sample no.: H2			
Theoretical value:		0.600		Theoretical value:		6.000	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		27		Number of laboratories:		28	
Arithmetic mean value:		0.618		Arithmetic mean value:		5.831	
Median:		0.624		Median:		5.985	
Standard deviation		0.163		Standard deviation		1.092	
Rel. st. deviation (%)		26.383		Rel. st. deviation (%)		18.723	
Run 2:				Run 2:			
Number of laboratories:		26		Number of laboratories:		27	
Arithmetic mean value:		0.642		Arithmetic mean value:		6.026	
Median:		0.629		Median:		6.000	
Standard deviation		0.109		Standard deviation		0.366	
Rel. st. deviation (%)		16.914		Rel. st. deviation (%)		6.079	
Results in decreasing order:				Results in decreasing order:			
178	0.826	23	0.623	178	6.851	23	5.970
24	0.800	110	0.600	33	6.700	121	5.970
32	0.800	169	0.581	125	6.540	168	5.920
33	0.800	15	0.580	2	6.397	14	5.904
159	0.780	14	0.577	179	6.360	169	5.860
179	0.770	8	0.568	159	6.300	32	5.800
115	0.766	168	0.560	171	6.270	127	5.800
38	0.697	31	0.550	15	6.219	31	5.770
36	0.650	161	0.550	36	6.200	4	5.700
171	0.641	5	0.541	24	6.100	8	5.690
121	0.640	127	0.500	115	6.080	5	5.554
125	0.640	120	0.400	3	6.020	118	5.500
2	0.634	118	0.000 (*)	161	6.020	120	5.200
3	0.624	4 <	1.500	110	6.000	38	0.571 (*)
Arsenic				Arsenic			
Sample no.: H3				Sample no.: H4			
Theoretical value:		5.000		Theoretical value:		0.400	
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:		28		Number of laboratories:		27	
Arithmetic mean value:		5.009		Arithmetic mean value:		0.421	
Median:		4.930		Median:		0.422	
Standard deviation		0.404		Standard deviation		0.148	
Rel. st. deviation (%)		8.063		Rel. st. deviation (%)		35.194	
Run 2:				Run 2:			
Number of laboratories:		27		Number of laboratories:		24	
Arithmetic mean value:		4.972		Arithmetic mean value:		0.436	
Median:		4.930		Median:		0.426	
Standard deviation		0.360		Standard deviation		0.078	
Rel. st. deviation (%)		7.237		Rel. st. deviation (%)		17.968	
Results in decreasing order:				Results in decreasing order:			
178	6.010 (*)	121	4.930	33	0.800 (*)	121	0.420
33	5.700	24	4.900	159	0.620	3	0.401
125	5.650	169	4.900	24	0.600	110	0.400
179	5.650	23	4.870	115	0.550	127	0.400
2	5.355	31	4.840	178	0.534	169	0.392
159	5.300	168	4.840	38	0.492	14	0.385
171	5.216	4	4.800	23	0.469	8	0.381
115	5.200	127	4.800	125	0.451	5	0.361
36	5.190	8	4.760	179	0.450	31	0.360
3	5.040	5	4.639	36	0.448	161	0.330
110	5.040	120	4.500	2	0.435	32	0.300
15	5.037	32	4.400	15	0.433	120	0.100 (*)
14	4.971	118	4.400	168	0.430	118	0.000 (*)
38	4.930	161	4.390	171	0.422	4 <	1.500

Table 20: Analytical results for Cd in synthetic precipitation samples, 2010.

Cadmium				Cadmium			
Sample no.: H1				Sample no.: H2			
Theoretical value: 0.060				Theoretical value: 0.800			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 31				Number of laboratories: 36			
Arithmetic mean value: 0.106				Arithmetic mean value: 0.795			
Median: 0.060				Median: 0.790			
Standard deviation 0.230				Standard deviation 0.183			
Rel. st. deviation (%) 216.995				Rel. st. deviation (%) 22.997			
Run 2:				Run 2:			
Number of laboratories: 30				Number of laboratories: 35			
Arithmetic mean value: 0.065				Arithmetic mean value: 0.769			
Median: 0.060				Median: 0.790			
Standard deviation 0.021				Standard deviation 0.098			
Rel. st. deviation (%) 31.810				Rel. st. deviation (%) 12.749			
Results in decreasing order:				Results in decreasing order:			
129	1.340 (*)	16	0.060	129	1.700 (*)	159	0.790
178	0.133	31	0.060	178	0.904	169	0.788
24	0.120	112	0.060	32	0.900	15	0.786
3	0.094	121	0.060	121	0.890	33	0.780
36	0.080	161	0.060	171	0.874	118	0.780
174	0.075	8	0.059	3	0.857	23	0.757
2	0.070	169	0.058	2	0.843	8	0.756
13	0.070	5	0.055	10	0.830	31	0.750
179	0.070	168	0.055	125	0.828	110	0.750
23	0.068	141	0.048	168	0.828	5	0.745
171	0.068	10	0.045	115	0.821	141	0.740
14	0.065	176	0.045	36	0.820	161	0.740
115	0.064	32	0.040	112	0.820	174	0.710
38	0.062	33	0.040	16	0.800	4	0.690
125	0.061	118	0.040	120	0.800	127	0.600
15	0.060	132	< 10.000	14	0.797	13	0.550
		4	< 0.350	38	0.791	176	0.510
		127	< 0.200	24	0.790	179	0.500
		110	< 0.100			132	< 10.000
		159	< 0.010				
Cadmium				Cadmium			
Sample no.: H3				Sample no.: H4			
Theoretical value: 0.500				Theoretical value: 0.050			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 35				Number of laboratories: 30			
Arithmetic mean value: 0.489				Arithmetic mean value: 0.057			
Median: 0.490				Median: 0.050			
Standard deviation 0.055				Standard deviation 0.037			
Rel. st. deviation (%) 11.215				Rel. st. deviation (%) 64.065			
Run 2:				Run 2:			
Number of laboratories: 33				Number of laboratories: 29			
Arithmetic mean value: 0.490				Arithmetic mean value: 0.051			
Median: 0.490				Median: 0.050			
Standard deviation 0.041				Standard deviation 0.012			
Rel. st. deviation (%) 8.369				Rel. st. deviation (%) 24.250			
Results in decreasing order:				Results in decreasing order:			
178	0.630 (*)	159	0.490	24	0.240 (*)	16	0.050
179	0.570	15	0.486	178	0.092	31	0.050
171	0.548	24	0.480	174	0.068	112	0.050
3	0.538	110	0.480	23	0.065	161	0.050
125	0.534	174	0.480	36	0.065	179	0.050
168	0.529	169	0.476	13	0.060	169	0.048
36	0.527	8	0.475	32	0.060	8	0.047
14	0.526	31	0.470	3	0.056	5	0.042
115	0.524	5	0.461	2	0.053	14	0.040
121	0.520	23	0.459	38	0.053	33	0.040
10	0.519	161	0.450	115	0.053	121	0.040
2	0.515	141	0.448	125	0.053	141	0.038
112	0.510	13	0.430	171	0.053	176	0.034
38	0.509	33	0.420	168	0.051	10	0.033
16	0.500	4	0.400	15	0.050	118	0.030
32	0.500	127	0.400			132	< 10.000
120	0.500	176	0.320 (*)			129	< 5.000
118	0.490	132	< 10.000			4	< 0.350
		129	< 5.000			127	< 0.200
						110	< 0.100
						159	< 0.010




Figure.9: Youden plot of nickel, 2010.




Figure.10: Youden plot of copper, 2010.




Figure.11: Youden plot of zinc, 2010.




Figure.12: Youden plot of arsenic, 2010.




Figure 14: Youden plot of lead, 2010.