


The twenty-second intercomparison of analytical methods within EMEP

Anne-Gunn Hjellbrekke, Hilde Th. Uggerud,
Jan Erik Hanssen and Jan Schaug



NILU : EMEP/CCC-Report 8/2005
REFERENCE : O-7729
DATE : SEPTEMBER 2005

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**The twenty-second intercomparison of
analytical methods within EMEP**

**Anne-Gunn Hjellbrekke, Hilde Th. Uggerud,
Jan Erik Hanssen and Jan Schaug**



Norwegian Institute for Air Research
P.O. Box 100, N-2027 Kjeller, Norway

Contents

	Page
1. Introduction.....	5
2. Organisation of the intercomparison	5
3. Data handling	6
3.1 Data analysis.....	6
3.2 Bar plots.....	6
3.3 Youden plot	6
3.4 Estimating random and systematic errors from laboratory comparisons.....	8
3.4.1 Estimating random errors	8
3.4.2 Estimating systematic errors.....	9
4. Results	10
4.1 Sulphur dioxide in absorbing solution (A-samples).....	10
4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	10
4.3 Nitrogen dioxide in absorbing solution (C-samples).....	10
4.4 Ammonia on impregnated filters (J-samples)	11
4.5 Precipitation (G-samples).....	11
4.5.1 Conductivity and ion balance	11
5. Summary.....	12
6. References.....	12
Appendix 1 Tables	15
Appendix 2 Figures	49

The twenty-second intercomparison of analytical methods within EMEP

1. Introduction

36 different laboratories in European countries are performing chemical analysis of air and precipitation samples within EMEP (Co-operative Programme for Monitoring and Evaluation of Long-range Transmission of Air Pollutants in Europe). Since the measurement programme is based on individual national networks, the participating laboratories apply different sampling and analytical methods. Most of the methods used are described in the manual for sampling and chemical analysis (EMEP, 1996).

In order to improve the data comparability and to get a picture of the different laboratories' performance, interlaboratory comparisons are organised by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research (NILU). So far twenty-two intercomparisons have been arranged (Hanssen, 1988, 1990; Hanssen et al., 1983; Hanssen and Ladegård, 1984, 1985, 1987; Hanssen and Skjelmoen, 1992, 1994, 1995, 1996, 1997, 2001; Thrane, 1978, 1980a, 1980b, 1981; Uggerud et al., 2001, 2002, 2003, 2004).

Since 2000 the laboratory intercomparisons within EMEP have also been open for participation of laboratories from other networks.

This report gives the results of the twenty-second interlaboratory test.

2. Organisation of the intercomparison

The samples for the twenty-second intercomparison (see Table 2) were prepared and distributed to 68 laboratories in July 2004.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 15 October 2004. A total of 68 laboratories have returned their results. This includes 32 EMEP-laboratories.

The participating laboratories received the theoretical (expected) values by e-mail 26.11.2004. The laboratories were given the opportunity to compare their results with the expected ones, and give corrected values if obvious mistakes e.g. misprints had occurred. A few corrections were reported. In those cases the corrected values are used in this report. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 4–8.

3. Data handling

The data reported from the participants are presented in Tables 9, 11, 13, 15, 17 and 19–29.

3.1 Data analysis

The reported values are presented in the tables in decreasing order together with the number of the laboratory. The expected (theoretical) value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation in percent are also given. After the first statistical run with all results included, the calculation was repeated with the outliers excluded. The outliers (unused) are defined as the results more than two standard deviations from the mean value in the first run.

The ratio between expected values (theoretical) to reported values, the ratio between measured to calculated conductivity and the ratio between equivalent concentrations of anions to equivalent concentrations cations, are presented in tables.

3.2 Bar plots

Bar-plots are used for the graphical presentation of the data. Figures 2–16 are showing the relative deviation from expected value for the different laboratories. There is one plot for each single sample.

Figure 17 gives median compared to expected value for the results reported by EMEP-laboratories and the other participating laboratories, respectively.

3.3 Youden plot

The Youden plot is a graphical method to analyse inter-laboratory data where the samples are ordered in pairs with similar concentrations. One plot is made for each pair of samples and gives results for all participating laboratories. The plots visualize both systematic and random errors.

The plot is drawn as a scatter plot where each point represents a pair of concentrations for one laboratory. The expected values for the two samples are drawn as solid blue lines. The arithmetic average of the measured values excluding outliers are drawn as dotted lines. The solid lines divide the plot in four quadrants and a 45° reference line going through the intercept of the solid lines may be added.

If errors are due to random factors, the points will be evenly distributed around the mean value and situated in all four quadrants.

If systematic errors dominate, the results will be close to 45° reference line, and be situated in the upper right quadrant (overestimation) or lower left quadrant (underestimation).

Drawing a line from a given point perpendicular on the 45° reference line gives two line segments, one from the point to the intercept on the reference line (a),

and one continuing from the intercept to the point representing the expected values (b). The lengths of these line segments are measures of the random and systematic errors respectively.

Ellipses with radii corresponding to the data quality objectives (DQO, Table 1) are added in each plot. The data points are colour coded depending on the magnitude of errors as given in Table 1.




Figure 1: Youden plot showing concentrations for a pair of samples (green), expected values (blue lines), average of measured values (dotted lines) and random and systematic errors (red lines)

In Figures 18–33 the reported data are presented in Youden plots.

Table 1: Youden plot parameters.


Radii	Components
10%	SO ₂ in abs.sol, NO ₂ in abs.sol.
20%	SO ₂ , HNO ₃ and NH ₃ in impregnated filter
Radii = DQO	Components
10% accuracy or better	SO ₄ ²⁻ , NO ₃ ⁻
15% accuracy or better	NH ₄ ⁺ , Cl ⁻ , Ca ²⁺ , K ⁺ , Mg ²⁺ , Na ⁺ , cond, H ⁺ (from pH)
0.1 units	pH
Criteria	Colour
Within 0.5*DQO	Blue
Within DQO	Green
Within 2*DQO	Orange
> 2*DQO	Red

3.4 Estimating random and systematic errors from laboratory comparisons

Table 34 presents relative random and relative systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples. The calculation method and assumptions used are given in Chapter 3.4.1 and Chapter 3.4.2.

3.4.1 Estimating random errors

Systematic errors or bias in the laboratory analyses give a constant shift in the results from the expected ones at a particular concentration level. It is assumed that laboratories taking part in comparisons will obtain results near the expected ones when this bias is removed, and that the differences between expected and obtained results more often will be close to zero than not. A triangular distribution, based upon this assumption, can be used to quantify the random errors in the laboratory results (Eurachem/CITAC, 2000).



The triangle distribution is symmetric with a baseline $2a$. The height in the triangle will be $1/a$ when the triangle area equals 1. The standard uncertainty is given by

$$u(x) = \frac{a}{\sqrt{6}} \quad (1)$$

and more than 95 % of the data will be within $\pm 2 \cdot u(x)$. The distance from $-a$ to a (i.e. $2a$) is called the range. When applied on the laboratory comparison results, the range equals the distance between the largest and smallest of the four differences between expected and found concentrations. As long as the bias can be assumed to be constant for the samples in the comparison of a specific component, it cannot have an effect on the distance corresponding to $2a$. The bias may be dependent upon the concentrations, but can be considered approximate constant for the concentrations used here in the comparison of the main components in precipitation, since the differences between the concentrations are small.

L and T represent the laboratories' and the expected concentrations respectively, and D is the difference. The difference for the lowest concentration is

$$D_1 = L_1 - T_1 \quad (2)$$

and the differences are D_1, D_2, D_3, D_4 in increasing order.

The range is $D_4 - D_1$ and the standard uncertainty for the differences $u(D)$ becomes

$$u(D) = \frac{(D_4 - D_1)}{(2 \cdot \sqrt{6})}. \quad (3)$$

The average expected concentration T for the four samples is given by

$$T = \frac{(T_1 + T_2 + T_3 + T_4)}{4} \quad (4)$$

The relative standard uncertainty, RSD, for 4 samples is given by $\frac{u(D)}{T}$, or

$$RSD = \frac{2 \cdot (D_4 - D_1 \cdot 100)}{\sqrt{6} \cdot (T_1 + T_2 + T_3 + T_4)} \%, \quad (5)$$

and 95 per cent of the laboratory results in this comparison are expected to be within $\pm 2 \cdot RSD$.

If the data quality objectives (DQO) likewise are looked upon as 95 percentiles, then 95 per cent of the laboratory analytical results should not be more than 10 or 15 per cent from the correct values (10 per cent for S and N containing components and 15 per cent for other components).

Correspondingly, the values $2 \cdot RSD$ should therefore be less than 10 or 15 per cent in order to comply with the DQO.

3.4.2 Estimating systematic errors

An estimation of bias in single measurements requires a long data series, and four samples as we normally have in laboratory comparison, are merely able to give an indication of the bias or a very coarse estimate.

Coarse estimates have been performed here in the cases where the four samples had similar concentrations and where all four laboratory results were either higher or lower than the expected concentrations. The median of the differences D_i , as defined above, was taken as a measure of the bias, B , in these cases.

$$B = \text{median}[D_i] \quad (6)$$

A relative bias, RB , was also calculated based upon the average expected concentration T , as defined in (4).

$$RB = \frac{4 \cdot \text{median}[D_i] \cdot 100}{(T_1 + T_2 + T_3 + T_4)} \% \quad (7)$$

4. Results

4.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. The results are given in Table 9, Figures 2 and 18. For those laboratories that reported a blank value this has been subtracted from the reported results. The ratios of measured value to expected value are presented in Table 10.

The sulphate concentration in the sample solutions correspond to a SO₂ concentration in air of 3.11–6.63 µg S m⁻³, when 70 ml absorbing solution and 3.6 m³ sampling volume is used.

Only 5 laboratories have reported values for SO₂ in absorbing solution.

4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were analysed by 19 laboratories. The value reported for the blank filter was subtracted from the other values before the data were used.

The amount of sulphur on the distributed filters corresponds to air concentrations between 0.64–2.40 µg S m⁻³ when 25 m³ is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of HNO₃-N. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between 0.39 µg N m⁻³–1.51 µg N m⁻³ when 25 m³ sampling volume is used.

Sulphur dioxide results show prevalence of systematic versus random errors. The systematic error is most clearly for the low concentration samples.

Nitric acid on impregnated filters shows good agreement with expected values. Few outliers are reported and most results are within the 10% of expected value.

The results are presented in Tables 11 and 13 and Figures 3, 4, 19 and 20.

4.3 Nitrogen dioxide in absorbing solution (C-samples)

The four samples distributed were made to represent both absorption solutions and extracts from iodide-impregnated glass filters. The samples contain known amounts of sodium nitrite diluted in water. In order to assure sample stability and to give the laboratories the opportunity to use the matrix they use in their daily routine, the distributed samples were to be diluted 1:10. The results should be reported as the diluted concentrations.

The 10 times diluted samples correspond to air concentrations between 3.55–7.10 µg NO₂-N m⁻³, when 70 ml absorbing solution and 1.4 m³ are used. If

4 ml extraction solution and 0.7 m³ sampling volume are used, the samples correspond to air concentrations between 0.23-0.81 µg NO₂-N m⁻³.

Nitrogen dioxide in absorbing solution shows good agreement with expected values. Few outliers are reported and most results are within the 10% of expected value. The Youden plots show prevalence of systematic error versus random error.

The results are presented in Table 15 and Figures 5 and 21.

4.4 Ammonia on impregnated filters (J-samples)

For the second time impregnated filters for determination of ammonia were distributed. Six impregnated filters inclusive two unidentified blank filters were sent to 21 laboratories. 19 laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 17 and Figures 6 and 22.

The amount of nitrogen on the filters correspond to air concentrations between 0.30-1.52 µg N m⁻³, if 25 m³ sampling volume is used.

Several reported values are more than 20% away from expected value. Youden plot shows that for the low concentration filters a considerable number of random errors are present. For the high concentration filters several results are biased low.

4.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2629 single results from 68 laboratories were reported. 121 results were identified as outliers. This is ~4% of the data, which is about the same as obtained last year. It should be noted that 44% of the outliers are caused by only five laboratories. The results are presented in Tables 19–29 and Figures 7–16 and 23–33.

4.5.1 Conductivity and ion balance

In EMEP, conductivity measurements are mainly used for quality control reasons. When all the main ions in the precipitation have been measured, conductivity values are compared with values calculated from the reported results. Table 30 gives the ratios of the measured to the calculated values.

Low concentration ions do not contribute much to the sum of ionic conductivities. By looking at the ratio of measured to calculated conductivity, errors in determination of low concentration ions may not be revealed. To include low concentration ions in the quality control, ion balance control must be used. This ratio should be used as a tool in the quality control system for those laboratories that measure all main components. The ratios of equivalent concentrations of anions versus equivalent concentrations of cations are shown in Table 31.

The Youden plot of conductivity shows mainly systematic error, which may be due to bad calibration of the instrument.

5. Summary

A total of 68 laboratories participated in the twenty-second intercomparison. 32 of these laboratories are within the EMEP network.

For all the samples analysed, the deviations from theoretical values are calculated. Figure 17 shows the median values compared to the expected values for all the parameters. For the EMEP laboratories the median deviations for both low- and high concentration samples are less than 3% and for other participants less than 6%. This is slightly better than obtained in earlier intercomparisons..

As in earlier intercomparisons, outliers are defined as values that deviate more than two standard deviations from the mean value. Outliers occur for all samples and almost all parameters. Out of a total of 3016 single results, 139 are defined as outliers. This is 5% of the reported data, which is comparable to earlier intercomparisons.

In Table 32 the ratio of the median values to the theoretical values for all the parameters is presented. As can be seen from this table, all parameters have median values that are in good agreement with the theoretical values.

6. References

Eurachem/CITAC (2000) Quantifying uncertainty in analytical measurements. 2nd ed. URL: <http://www.measurementuncertainty.org/mu/QUAM2000-1.pdf>.

EMEP (1996) Manual for sampling and chemical analysis. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/95) (Newest revision in 2001).

Hanssen, J.E. (1988) The tenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 3/88).

Hanssen, J.E. (1990) The eleventh intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/90).

Hanssen, J.E. and Ladegård, N.E. (1984) The seventh intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/84).

Hanssen, J.E. and Ladegård, N.E. (1985) The eighth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/85).

Hanssen, J.E. and Ladegård, N.E. (1987) The ninth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/87).

- Hanssen, J.E., Ladegård, N.E. and Thrane, K.E. (1983) The sixth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/83).
- Hanssen, J.E. and Skjelmoen, J.E. (1992) The twelfth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 7/92).
- Hanssen, J.E. and Skjelmoen, J.E. (1994) The thirteenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/94).
- Hanssen, J.E. and Skjelmoen, J.E. (1995) The fourteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 3/95).
- Hanssen, J.E. and Skjelmoen, J.E. (1996) The fifteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/96).
- Hanssen, J.E. and Skjelmoen, J.E. (1997) The sixteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/97).
- Hanssen, J.E. and Skjelmoen, J.E. (2001) The seventeenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 10/2001).
- Thrane, K.E. (1978) Report on the first intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/78).
- Thrane, K.E. (1980a) Report on the second and third intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/80).
- Thrane, K.E. (1980b) Report on the fourth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 5/80).
- Thrane, K.E. (1981) The fifth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/81).
- Uggerud, H.T., Hanssen, J.E. and Skjelmoen, J.E. (2001) The eighteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 11/2001).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2002) The nineteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/2002).

Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2003) The twentieth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2003).

Uggerud, H.T., Hanssen, J.E., Schaug, J. and Hjellbrekke, A.-G. (2004) The twenty-first intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2004).

Appendix 1

Tables

Table 2: Samples distributed for the twenty-second interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 3a: EMEP laboratories participating in the twenty-second laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Austria	(1)	Umweltbundesamt Zweigstelle Sud, Klagenfurt
Canada	(26)	Meteorological Service of Canada, Toronto
Croatia	(35)	Meteorological and Hydrological Service of Croatia
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(6)	Laboratories Wolff
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messtelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Iceland	(11)	Idntæknistofnun Islands (Technological Inst. of Iceland)
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Italy	(30)	Join Research centre, Ispra
Latvia	(33)	Air Pollution Observation Laboratory
Lithuania	(32)	Atmospheric Pollution Research Laboratory, Institute of Physics, Vilnius
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(39)	Environmental Monitoring Laboratory, Institute of Environmental Protection
Portugal	(17)	Laboratorio Santo Andre
Romania	(18)	Research and Engineering Institute for Environment
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia and Montenegro	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
Turkey	(34)	Refik Saydam Institute, Ankara
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United States of America	(27)	Illinois State Water Survey

Table 3b: Participating laboratories outside the EMEP network.

Germany	(102)	Ökologie-Zentrum Universität München
Germany	(104)	Hessige Landwirtschaftliche
Finland	(107)	The Finnish Forest Institute
Germany	(108)	Institut f. Bondenkunde und Standortlehre, Dredsten
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (NVF)
Germany	(113)	Landesforstanstalt Eberswalde, abt. Waldökologie
Italy	(114)	C.N.R. Istituto Italiano di Idrobiologia
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(117)	Sächsische Landesanstalt für Forsten, Graupa
Germany	(118)	Forstliche Versuchs- und Forschungsanstalt
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landesamt für Natur und Umwelt
Belgium	(124)	Laboratorium voor Bondenkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
China	(131)	Chongqing Institute of Environmental Science and Monitoring
Belarus	(133)	Institute for Problems of Natural Resources Use and Ecology
China	(138)	Guizhou Research Institute of Environmental Protection Science, Guiyang
Denmark	(139)	Danish Centre for Forest, Landscape and Planning, Hørsholm
Italy	(140)	C.N.R. Istituto di Ricerca sulle Acque
France	(142)	EPLD, Centre de Lagor
Finland	(145)	Tartu Environmental Research, Tartu
Luxembourg	(146)	Cellule de Recherche en Environment et Biotechnologies Public Research Center-Gabriel Lippmann
Austria	(147)	Institute of Pollution Research and Forest Chemistry
Netherlands	(148)	Alterra, Wageningen
Switzerland	(149)	WSL Zentrallabor, Birmensdorf
Belgium	(151)	Laboratoire de l'Unité des Eaux et Forêt (EFOR), Louvain-la-Neuve
Norway	(152)	Norwegian Forest Research Institute, Ås
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
Portugal	(154)	Laboratorio de Analise Instrumental, Acores
Greece	(156)	Laboratory of Forest Land and Biogeochemistry, Ilisia
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata

Table 4: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
1. Ion chromatography	6, 15, 19, 21, 23

Table 5: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
1. Spectrophotometry	16
2. Ion chromatography	3, 4, 5, 8, 11, 12, 13, 15, 20, 22, 31,32, 33, 34, 36, 38, 116, 131, 138, 158
3. Capillary Ion Analysis	39

Table 6: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Reduction to nitrite	16
2. Ion chromatography	3, 4, 5, 8, 11, 13, 15, 20, 22, 31,32, 33, 34, 36, 116,131, 138, 158
3. Capillary Ion Analysis	39

Table 7: Analytical method for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	3, 4, 8, 10, 16, 19,32, 33,34 39, 116,
2. FIA	11
3 Ion chromatography	5, 13,15, 20, 36, 131, 138, 158

Table 8: Analytical method used for NO₂ in absorbing solution (C).

Method	Laboratory
1. Spectrophotometry	3, 4, 8, 12, 15, 16, 19, 20, 22, 23, 31,32, 33, 34, 35, 38, 39, 131, 138
Ion chromatography	36

Table 9: Analytical results for sulphur dioxide in absorbing solution.

SO2-S in absorbing solution				SO2-S in absorbing solution			
Sample no.: 1				Sample no.: 3			
Theoretical value:		0.321		Theoretical value:		0.361	
Unit: µg S / ml				Unit: µg S / ml			
Run 1:				Run 1:			
Number of laboratories:		6		Number of laboratories:		6	
Arithmetic mean value:		0.307		Arithmetic mean value:		0.351	
Median:		0.329		Median:		0.365	
Standard deviation		0.081		Standard deviation		0.051	
Rel. st. deviation (%)		26.415		Rel. st. deviation (%)		14.625	
Run 2:				Run 2:			
Number of laboratories:		6		Number of laboratories:		6	
Arithmetic mean value:		0.307		Arithmetic mean value:		0.351	
Median:		0.329		Median:		0.365	
Standard deviation		0.081		Standard deviation		0.051	
Rel. st. deviation (%)		26.415		Rel. st. deviation (%)		14.625	
Results in decreasing order:				Results in decreasing order:			
17	0.390	21	0.324	17	0.394	15	0.362
15	0.333	6	0.311	23	0.381	6	0.349
23	0.333	19	0.151	21	0.368	19	0.251
SO2-S in absorbing solution				SO2-S in absorbing solution			
Sample no.: 4				Sample no.: 5			
Theoretical value:		0.168		Theoretical value:		0.140	
Unit: µg S / ml				Unit: µg S / ml			
Run 1:				Run 1:			
Number of laboratories:		6		Number of laboratories:		6	
Arithmetic mean value:		0.222		Arithmetic mean value:		0.145	
Median:		0.174		Median:		0.140	
Standard deviation		0.119		Standard deviation		0.015	
Rel. st. deviation (%)		53.545		Rel. st. deviation (%)		9.987	
Run 2:				Run 2:			
Number of laboratories:		5		Number of laboratories:		6	
Arithmetic mean value:		0.174		Arithmetic mean value:		0.145	
Median:		0.172		Median:		0.140	
Standard deviation		0.018		Standard deviation		0.015	
Rel. st. deviation (%)		10.487		Rel. st. deviation (%)		9.987	
Results in decreasing order:				Results in decreasing order:			
6	0.462 (*)	15	0.172	17	0.174	15	0.137
17	0.201	21	0.170	23	0.145	19	0.137
23	0.176	19	0.150	21	0.143	6	0.136

Table 10: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide in absorbing solutions.

Lab. No	Measured value / Expected				Average
	Sample No				
	A1	A2	A3	A5	
6	0.97	0.97	2.74	0.97	1.41
15	1.04	1.00	1.02	0.98	1.01
19	0.47	0.70	0.89	0.98	0.76
21	1.01	1.02	1.01	1.02	1.02
23	1.04	1.06	1.05	1.03	1.04

Table 11: Analytical results for sulphur dioxide in impregnated filter.

SO2-S on impregnated filter				SO2-S on impregnated filter			
Sample no.: 1				Sample no.: 2			
Theoretical value:		22.044		Theoretical value:		16.032	
Unit: ugS/filter				Unit: ugS/filter			
Run 1:				Run 1:			
Number of laboratories:		19		Number of laboratories:		19	
Arithmetic mean value:		22.507		Arithmetic mean value:		16.762	
Median:		21.610		Median:		16.400	
Standard deviation		2.617		Standard deviation		2.166	
Rel. st. deviation (%)		11.627		Rel. st. deviation (%)		12.921	
Run 2:				Run 2:			
Number of laboratories:		18		Number of laboratories:		18	
Arithmetic mean value:		22.043		Arithmetic mean value:		16.373	
Median:		21.605		Median:		16.285	
Standard deviation		1.709		Standard deviation		1.389	
Rel. st. deviation (%)		7.754		Rel. st. deviation (%)		8.483	
Results in decreasing order:				Results in decreasing order:			
138	30.857 (*)	16	21.600	138	23.756 (*)	20	16.170
32	26.960	131	21.560	131	19.280	39	15.910
38	24.800	3	21.490	16	19.000	32	15.860
33	22.919	34	21.067	8	17.747	34	15.389
5	22.727	22	20.800	5	17.355	4	15.300
8	22.717	4	20.700	33	17.343	116	15.100
15	22.460	158	20.470	15	16.760	38	14.900
36	22.400	20	20.300	3	16.650	158	14.870
31	22.397	116	19.800	31	16.484	22	14.200
39	21.610			36	16.400		
SO2-S on impregnated filter				SO2-S on impregnated filter			
Sample no.: 4				Sample no.: 5			
Theoretical value:		52.075		Theoretical value:		60.087	
Unit: ugS/filter				Unit: ugS/filter			
Run 1:				Run 1:			
Number of laboratories:		18		Number of laboratories:		18	
Arithmetic mean value:		53.522		Arithmetic mean value:		60.423	
Median:		52.463		Median:		59.825	
Standard deviation		4.497		Standard deviation		4.070	
Rel. st. deviation (%)		8.403		Rel. st. deviation (%)		6.736	
Run 2:				Run 2:			
Number of laboratories:		16		Number of laboratories:		17	
Arithmetic mean value:		52.119		Arithmetic mean value:		59.791	
Median:		52.059		Median:		59.819	
Standard deviation		1.876		Standard deviation		3.153	
Rel. st. deviation (%)		3.600		Rel. st. deviation (%)		5.274	
Results in decreasing order:				Results in decreasing order:			
36	66.700 (*)	31	52.087	138	71.182 (*)	31	59.819
138	62.791 (*)	131	52.030	36	67.400	38	59.400
33	54.930	8	51.789	33	64.120	8	58.653
39	53.810	22	51.500	5	61.626	116	58.600
5	53.793	20	51.470	32	61.580	4	58.200
16	53.620	158	51.370	39	61.310	158	58.170
32	53.610	4	51.200	34	60.892	22	57.900
116	53.300	38	49.000	16	60.320	20	54.760
34	52.839	15	47.560	131	59.830	15	53.860

Table 12: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide on impregnated filters. The reported results are corrected for blank value (B1).

Lab. No	Measured / Expected value				Average
	Sample No				
	B1	B2	B4	B5	
3	0.97	1.04		1.01	1.01
4	0.94	0.95	0.98	0.97	0.96
5	1.03	1.08	1.03	1.03	1.04
8	1.03	1.11	0.99	0.98	1.03
15	1.02	1.05	0.91	0.90	0.97
16	0.98	1.19	1.03	1.00	1.05
20	0.92	1.01	0.99	0.91	0.96
22	0.94	0.89	0.99	0.96	0.95
31	1.02	1.03	1.00	1.00	1.01
32	1.22	0.99	1.03	1.02	1.07
33	1.04	1.08	1.05	1.07	1.06
34	0.96	0.96	1.01	1.01	0.99
36	1.02	1.02	1.28	1.12	1.11
38	1.13	0.93	0.94	0.99	1.00
39	0.98	0.99	1.03	1.02	1.01
116	0.90	0.94	1.02	0.98	0.96
131	0.98	1.20	1.00	1.00	1.04
138	1.40	1.48	1.21	1.18	1.32
158	0.93	0.93	0.99	0.97	0.95

Table 13: Analytical results for nitric acid on impregnated filter.

HNO ₃ -N on impregnated filter				HNO ₃ -N on impregnated filter			
Sample no.: 1				Sample no.: 2			
Theoretical value:		9.816		Theoretical value:		13.088	
Unit: ugN/filter				Unit: ugN/filter			
Run 1:				Run 1:			
Number of laboratories:		17		Number of laboratories:		18	
Arithmetic mean value:		9.551		Arithmetic mean value:		12.914	
Median:		9.690		Median:		13.008	
Standard deviation		0.917		Standard deviation		0.927	
Rel. st. deviation (%)		9.601		Rel. st. deviation (%)		7.178	
Run 2:				Run 2:			
Number of laboratories:		16		Number of laboratories:		17	
Arithmetic mean value:		9.689		Arithmetic mean value:		13.053	
Median:		9.705		Median:		13.125	
Standard deviation		0.744		Standard deviation		0.739	
Rel. st. deviation (%)		7.677		Rel. st. deviation (%)		5.659	
Results in decreasing order:				Results in decreasing order:			
32	10.680	31	9.516	32	14.270	20	12.890
5	10.605	158	9.510	8	14.079	158	12.710
8	10.378	15	9.500	5	13.775	39	12.700
131	10.344	20	9.340	138	13.713	31	12.679
39	10.200	4	9.100	131	13.484	16	12.570
138	10.070	22	8.800	3	13.440	4	12.500
36	9.825	34	7.742	15	13.200	33	12.365
16	9.720	33	7.348 (*)	116	13.200	22	11.200
3	9.690	116	<0.300	36	13.125	34	10.558 (*)
HNO ₃ -N on impregnated filter				HNO ₃ -N on impregnated filter			
Sample no.: 4				Sample no.: 5			
Theoretical value:		34.356		Theoretical value:		37.628	
Unit: ugN/filter				Unit: ugN/filter			
Run 1:				Run 1:			
Number of laboratories:		17		Number of laboratories:		17	
Arithmetic mean value:		33.920		Arithmetic mean value:		36.382	
Median:		34.206		Median:		36.510	
Standard deviation		2.038		Standard deviation		2.208	
Rel. st. deviation (%)		6.010		Rel. st. deviation (%)		6.070	
Run 2:				Run 2:			
Number of laboratories:		16		Number of laboratories:		17	
Arithmetic mean value:		34.215		Arithmetic mean value:		36.382	
Median:		34.308		Median:		36.510	
Standard deviation		1.690		Standard deviation		2.208	
Rel. st. deviation (%)		4.938		Rel. st. deviation (%)		6.070	
Results in decreasing order:				Results in decreasing order:			
36	37.825	33	33.933	33	40.227	4	36.000
32	36.490	116	33.900	32	39.850	31	35.983
39	35.300	4	33.600	138	38.577	16	35.980
131	34.944	16	33.240	39	37.600	34	35.701
138	34.803	31	33.093	36	37.425	116	34.900
20	34.700	34	32.383	5	37.280	20	34.700
5	34.412	15	30.200	8	36.931	15	32.200
158	34.410	22	29.200 (*)	131	36.634	22	32.000
8	34.206			158	36.510		

Table 14: The ratios of the theoretical values and the results found by the laboratories in the determination of nitric acid on impregnated filters. The reported results are corrected for blank value (B1).

Lab.No	Measured / Expected value				Average
	Sample No				
	B1	B2	B4	B5	
3	0.99	1.03		0.99	1.01
4	0.93	0.96	0.98	0.96	0.96
5	1.08	1.05	1.00	0.99	1.01
8	1.06	1.08	1.00	0.98	1.02
15	0.97	1.01	0.88	0.86	0.91
16	0.99	0.96	0.97	0.96	0.96
20	0.95	0.98	1.01	0.92	0.97
22	0.90	0.86	0.85	0.85	0.85
31	0.97	0.97	0.96	0.96	0.96
32	1.09	1.09	1.06	1.06	1.07
33	0.75	0.94	0.99	1.07	1.00
34	0.79	0.81	0.94	0.95	0.90
36	1.00	1.00	1.10	0.99	1.03
39	1.04	0.97	1.03	1.00	1.00
116		1.03	1.00	0.94	0.99
131	1.05	1.03	1.02	0.97	1.01
138	1.03	1.048	1.01	1.03	1.03
158	0.97	0.971	1.00	0.97	0.98

Table 15: Analytical results for nitrogen dioxide in absorbing solution.

NO2-N in absorbing solution				NO2-N in absorbing solution			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.112				Theoretical value: 0.071			
Unit: ug N/ml				Unit: ug N/ml			
Run 1:				Run 1:			
Number of laboratories: 21				Number of laboratories: 21			
Arithmetic mean value: 0.126				Arithmetic mean value: 0.081			
Median: 0.113				Median: 0.071			
Standard deviation 0.060				Standard deviation 0.036			
Rel. st. deviation (%) 47.286				Rel. st. deviation (%) 44.703			
Run 2:				Run 2:			
Number of laboratories: 20				Number of laboratories: 20			
Arithmetic mean value: 0.114				Arithmetic mean value: 0.073			
Median: 0.113				Median: 0.071			
Standard deviation 0.014				Standard deviation 0.009			
Rel. st. deviation (%) 12.332				Rel. st. deviation (%) 12.555			
Results in decreasing order:				Results in decreasing order:			
20	0.380 (*)	3	0.112	20	0.233 (*)	131	0.071
19	0.166	8	0.111	19	0.109	32	0.071
38	0.120	31	0.110	38	0.077	33	0.070
12	0.118	33	0.110	12	0.077	138	0.070
34	0.116	36	0.109	10	0.076	23	0.070
32	0.116	4	0.108	22	0.075	36	0.070
39	0.116	131	0.107	31	0.073	35	0.069
22	0.115	15	0.105	8	0.072	4	0.068
10	0.115	138	0.100	3	0.071	15	0.067
23	0.114	16	0.091	34	0.071	16	0.062
35	0.113			39	0.071		
NO2-N in absorbing solution				NO2-N in absorbing solution			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.125				Theoretical value: 0.142			
Unit: ug N/ml				Unit: ug N/ml			
Run 1:				Run 1:			
Number of laboratories: 21				Number of laboratories: 21			
Arithmetic mean value: 0.145				Arithmetic mean value: 0.164			
Median: 0.127				Median: 0.145			
Standard deviation 0.074				Standard deviation 0.082			
Rel. st. deviation (%) 50.922				Rel. st. deviation (%) 50.092			
Run 2:				Run 2:			
Number of laboratories: 20				Number of laboratories: 20			
Arithmetic mean value: 0.129				Arithmetic mean value: 0.146			
Median: 0.127				Median: 0.144			
Standard deviation 0.015				Standard deviation 0.017			
Rel. st. deviation (%) 11.493				Rel. st. deviation (%) 11.736			
Results in decreasing order:				Results in decreasing order:			
20	0.461 (*)	36	0.127	20	0.515 (*)	23	0.144
19	0.186	39	0.127	19	0.211	36	0.144
38	0.137	33	0.126	38	0.160	3	0.143
32	0.133	3	0.126	10	0.150	39	0.142
10	0.132	31	0.125	22	0.150	33	0.140
34	0.131	131	0.123	34	0.149	131	0.140
22	0.130	4	0.122	35	0.149	4	0.137
35	0.130	138	0.120	32	0.148	15	0.133
12	0.130	15	0.119	31	0.146	138	0.130
8	0.127	16	0.106	8	0.145	16	0.123
23	0.127			12	0.145		

Table 16: The ratios of the theoretical values and the results found by the laboratories in the determination of nitrogen dioxide in absorbing solutions.

Lab.No	Measured / Expected				Average
	Sample No.				
	C1	C2	C3	C4	
3	1.00	1.00	1.01	1.01	1.00
4	0.97	0.96	0.97	0.96	0.97
8	0.99	1.02	1.01	1.02	1.01
10	1.03	1.07	1.05	1.06	1.05
12	1.06	1.08	1.04	1.02	1.05
15	0.94	0.94	0.95	0.94	0.94
16	0.81	0.87	0.85	0.87	0.85
19	1.49	1.53	1.49	1.48	1.50
20	1.04	1.00	1.12	1.10	1.07
22	1.03	1.06	1.04	1.06	1.04
23	1.02	0.98	1.01	1.01	1.01
31	0.98	1.02	1.00	1.03	1.01
32	1.04	1.00	1.06	1.04	1.04
33	0.98	0.98	1.01	0.98	0.99
34	1.04	1.00	1.05	1.05	1.03
35	1.01	0.97	1.04	1.05	1.02
36	0.98	0.98	1.01	1.01	1.00
38	1.07	1.08	1.09	1.13	1.09
39	1.04	1.00	1.01	1.00	1.01
131	0.96	0.99	0.98	0.98	0.98
138	0.90	0.98	0.96	0.91	0.94

Table 17: Analytical results for ammonia on impregnated filter.

NH3-N on impregnated filter
 Sample no.: 1
 Theoretical value: 14.035
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 13.983
 Median: 13.675
 Standard deviation 1.535
 Rel. st. deviation (%) 10.977

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 13.753
 Median: 13.648
 Standard deviation 1.198
 Rel. st. deviation (%) 8.708

Results in decreasing order:

33	18.115 (*)	32	13.620
13	15.610	5	13.450
131	15.595	20	13.297
8	15.531	158	13.235
116	14.615	39	13.140
138	14.356	34	13.005
16	14.335	19	12.680
15	14.260	3	12.310
4	13.905	10	10.943
36	13.675		

NH3-N on impregnated filter
 Sample no.: 4
 Theoretical value: 8.521
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 8.290
 Median: 8.250
 Standard deviation 0.778
 Rel. st. deviation (%) 9.381

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 8.178
 Median: 8.225
 Standard deviation 0.622
 Rel. st. deviation (%) 7.609

Results in decreasing order:

13	10.310 (*)	5	8.200
131	9.325	158	8.155
116	9.175	39	8.080
138	8.677	16	7.855
36	8.655	34	7.815
4	8.605	19	7.440
8	8.495	20	7.394
32	8.370	3	7.270
33	8.365	10	7.080
15	8.250		

NH3-N on impregnated filter
 Sample no.: 3
 Theoretical value: 22.055
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 21.579
 Median: 21.705
 Standard deviation 1.874
 Rel. st. deviation (%) 8.685

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 21.590
 Median: 21.705
 Standard deviation 1.391
 Rel. st. deviation (%) 6.444

Results in decreasing order:

13	25.500 (*)	15	21.560
8	25.022	3	21.090
131	23.475	39	20.940
116	22.585	158	20.935
138	22.444	16	20.595
32	21.960	20	20.277
36	21.905	19	19.460
5	21.890	34	19.320
33	21.865	10	17.472 (*)
4	21.705		

NH3-N on impregnated filter
 Sample no.: 5
 Theoretical value: 38.095
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 36.787
 Median: 37.235
 Standard deviation 4.686
 Rel. st. deviation (%) 12.737

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 37.502
 Median: 37.350
 Standard deviation 3.602
 Rel. st. deviation (%) 9.605

Results in decreasing order:

13	45.980	5	36.390
32	42.020	3	36.310
8	41.681	158	36.235
4	39.005	19	35.060
131	38.705	20	34.857
15	38.660	34	34.653
138	38.595	39	34.540
116	38.475	10	29.168
33	37.465	16	23.925 (*)
36	37.235		

Table 18: The ratios of the theoretical values and the results found by the laboratories in the determination of ammonia on impregnated filters. The reported results are corrected for an average blank value (J1 and J6).

Lab. no.	Measured / Expected				Average
	Sample No.				
	J1	J3	J4	J5	
3	0.88	0.96	0.85	0.95	0.91
4	0.99	0.98	1.01	1.02	1.00
5	0.96	0.99	0.96	0.96	0.97
8	1.11	1.13	1.00	1.09	1.08
10	0.78	0.79	0.83	0.77	0.79
13	1.11	1.16	1.21	1.21	1.17
15	1.02	0.98	0.97	1.01	0.99
16	1.02	0.93	0.92	0.63	0.88
19	0.90	0.88	0.87	0.92	0.89
20	0.95	0.92	0.87	0.91	0.91
32	0.97	1.00	0.98	1.10	1.01
33	1.29	0.99	0.98	0.98	1.06
34	0.93	0.88	0.92	0.91	0.91
36	0.97	0.99	1.02	0.98	0.99
39	0.94	0.95	0.95	0.91	0.94
116	1.04	1.02	1.08	1.01	1.04
131	1.11	1.06	1.09	1.02	1.07
138	1.02	1.02	1.02	1.01	1.02
158	0.94	0.95	0.96	0.95	0.95

Table 19: Analytical results for sulphate in precipitations samples.

Sulphate in precipitation				Sulphate in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 1.620				Theoretical value: 1.764			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 1.607				Arithmetic mean value: 1.781			
Median: 1.613				Median: 1.767			
Standard deviation 0.305				Standard deviation 0.359			
Rel. st. deviation (%) 18.974				Rel. st. deviation (%) 20.144			
Run 2:				Run 2:			
Number of laboratories: 63				Number of laboratories: 64			
Arithmetic mean value: 1.614				Arithmetic mean value: 1.772			
Median: 1.619				Median: 1.768			
Standard deviation 0.116				Standard deviation 0.109			
Rel. st. deviation (%) 7.163				Rel. st. deviation (%) 6.141			
Results in decreasing order:				Results in decreasing order:			
22	3.464 (*)	131	1.605	22	4.229 (*)	32	1.767
133	1.939	124	1.600	133	2.061	1	1.760
148	1.871	142	1.600	148	2.021	152	1.760
118	1.810	152	1.600	118	2.010	131	1.759
30	1.794	102	1.600	30	1.975	115	1.750
139	1.780	110	1.600	154	1.910	153	1.750
39	1.749	158	1.600	35	1.887	113	1.750
7	1.733	15	1.595	116	1.886	126	1.748
154	1.720	23	1.593	107	1.878	158	1.740
107	1.717	6	1.580	10	1.866	108	1.740
35	1.715	26	1.578	125	1.850	23	1.734
10	1.707	20	1.573	140	1.850	110	1.730
13	1.694	146	1.570	139	1.840	26	1.725
125	1.680	36	1.564	7	1.835	11	1.722
140	1.680	16	1.563	13	1.832	6	1.720
14	1.670	114	1.560	14	1.830	114	1.720
33	1.667	38	1.550	39	1.829	20	1.714
24	1.660	153	1.540	24	1.820	16	1.713
21	1.655	147	1.540	33	1.813	36	1.711
4	1.652	32	1.532	21	1.801	146	1.710
5	1.652	113	1.530	12	1.800	147	1.700
12	1.650	117	1.520	124	1.800	117	1.690
149	1.648	112	1.510	4	1.797	31	1.686
116	1.647	120	1.502	149	1.794	156	1.670
138	1.643	121	1.472	3	1.794	120	1.666
27	1.638	31	1.464	138	1.793	121	1.666
8	1.635	19	1.451	27	1.784	38	1.660
156	1.630	126	1.435	142	1.780	112	1.640
3	1.627	18	1.324	5	1.780	18	1.630
11	1.624	145	1.179	102	1.780	19	1.584
115	1.620	151	0.990 (*)	8	1.777	145	1.295
1	1.620	108	0.900 (*)	15	1.775	151	1.060 (*)
34	1.619	104	0.620 (*)	17	1.768	104	0.670 (*)
17	1.613			34	1.767		
Sulphate in precipitation				Sulphate in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.897				Theoretical value: 0.984			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 0.897				Arithmetic mean value: 0.985			
Median: 0.900				Median: 0.980			
Standard deviation 0.172				Standard deviation 0.188			
Rel. st. deviation (%) 19.164				Rel. st. deviation (%) 19.075			
Run 2:				Run 2:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 0.899				Arithmetic mean value: 0.985			
Median: 0.900				Median: 0.981			
Standard deviation 0.065				Standard deviation 0.082			
Rel. st. deviation (%) 7.204				Rel. st. deviation (%) 8.307			
Results in decreasing order:				Results in decreasing order:			
22	1.927 (*)	110	0.900	22	2.106 (*)	153	0.980
148	1.089	17	0.898	10	1.273	154	0.980
133	1.009	1	0.895	148	1.191	3	0.980
10	0.997	16	0.893	116	1.120	20	0.978
118	0.990	36	0.891	118	1.090	17	0.977
116	0.986	153	0.890	133	1.072	121	0.971
102	0.980	20	0.890	30	1.067	115	0.970
142	0.980	115	0.880	13	1.060	139	0.970
30	0.979	114	0.880	102	1.050	16	0.970
13	0.963	131	0.880	142	1.050	34	0.969
19	0.957	26	0.878	107	1.038	36	0.968
14	0.950	121	0.875	35	1.031	23	0.963
33	0.948	108	0.875	14	1.030	114	0.960
156	0.945	11	0.874	39	1.029	6	0.960
35	0.940	6	0.870	7	1.028	26	0.958
4	0.934	15	0.870	140	1.020	131	0.957
7	0.930	113	0.870	19	1.017	126	0.955
154	0.930	158	0.870	4	1.016	158	0.950
140	0.920	126	0.868	138	1.012	11	0.950
12	0.920	23	0.868	156	1.010	117	0.950
138	0.919	24	0.860	33	1.010	152	0.950
39	0.919	38	0.860	12	1.010	146	0.950
21	0.915	152	0.860	5	1.002	32	0.931
107	0.912	120	0.852	149	1.001	147	0.930
8	0.909	147	0.850	21	1.000	120	0.922
3	0.909	146	0.830	124	1.000	38	0.910
27	0.906	32	0.814	8	0.996	112	0.900
149	0.905	112	0.810	125	0.995	31	0.855
5	0.904	31	0.757	27	0.991	145	0.810
124	0.900	145	0.703	110	0.990	108	0.785
125	0.900	151	0.670	24	0.990	151	0.670
117	0.900	104	0.370 (*)	15	0.985	18	0.429 (*)
139	0.900	18	0.306 (*)	1	0.982	104	0.400 (*)
34	0.900			113	0.980		

Table 20: Analytical results for nitrate in precipitations samples.

Nitrate in precipitation				Nitrate in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.442				Theoretical value: 0.376			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 66			
Arithmetic mean value: 0.438				Arithmetic mean value: 0.372			
Median: 0.439				Median: 0.373			
Standard deviation 0.047				Standard deviation 0.031			
Rel. st. deviation (%) 10.720				Rel. st. deviation (%) 8.302			
Run 2:				Run 2:			
Number of laboratories: 62				Number of laboratories: 61			
Arithmetic mean value: 0.435				Arithmetic mean value: 0.375			
Median: 0.439				Median: 0.374			
Standard deviation 0.023				Standard deviation 0.019			
Rel. st. deviation (%) 5.351				Rel. st. deviation (%) 5.059			
Results in decreasing order:				Results in decreasing order:			
40	0.618 (*)	11	0.438	40	0.487 (*)	116	0.373
102	0.590 (*)	26	0.438	142	0.430	146	0.373
142	0.590 (*)	3	0.434	102	0.430	16	0.372
19	0.510	22	0.432	19	0.428	12	0.370
39	0.477	1	0.430	108	0.402	118	0.370
108	0.476	6	0.430	124	0.400	1	0.370
23	0.463	24	0.430	154	0.400	6	0.370
140	0.460	36	0.430	31	0.395	104	0.370
139	0.460	38	0.430	30	0.393	21	0.369
154	0.460	110	0.430	140	0.390	138	0.368
30	0.454	113	0.430	15	0.390	3	0.368
125	0.453	115	0.430	120	0.390	36	0.367
149	0.451	158	0.430	39	0.389	33	0.362
32	0.451	33	0.429	23	0.388	34	0.360
114	0.450	34	0.427	13	0.385	139	0.360
15	0.450	35	0.423	17	0.384	115	0.360
5	0.450	116	0.423	5	0.384	126	0.360
31	0.448	4	0.421	125	0.383	24	0.360
8	0.446	104	0.420	121	0.383	158	0.360
131	0.446	117	0.420	149	0.382	4	0.357
121	0.446	107	0.411	32	0.382	35	0.354
148	0.445	14	0.410	113	0.380	147	0.350
17	0.445	112	0.410	114	0.380	14	0.350
10	0.445	153	0.410	131	0.379	110	0.350
7	0.444	13	0.410	8	0.379	153	0.350
146	0.443	147	0.400	7	0.378	38	0.350
20	0.442	152	0.400	26	0.378	117	0.350
27	0.441	156	0.400	20	0.377	107	0.346
118	0.440	124	0.400	148	0.377	152	0.340
16	0.440	126	0.380	10	0.376	151	0.310 (*)
120	0.440	133	0.358	22	0.376	145	0.298 (*)
12	0.440	151	0.330 (*)	27	0.374	133	0.282 (*)
138	0.440	145	0.290 (*)	11	0.374	156	0.277 (*)
		21	<0.439				
		112	<0.370				
Nitrate in precipitation				Nitrate in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.663				Theoretical value: 0.590			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 0.654				Arithmetic mean value: 0.582			
Median: 0.652				Median: 0.587			
Standard deviation 0.057				Standard deviation 0.040			
Rel. st. deviation (%) 8.665				Rel. st. deviation (%) 6.864			
Run 2:				Run 2:			
Number of laboratories: 63				Number of laboratories: 64			
Arithmetic mean value: 0.655				Arithmetic mean value: 0.582			
Median: 0.652				Median: 0.587			
Standard deviation 0.027				Standard deviation 0.025			
Rel. st. deviation (%) 4.172				Rel. st. deviation (%) 4.312			
Results in decreasing order:				Results in decreasing order:			
133	0.861 (*)	1	0.650	40	0.711 (*)	148	0.587
40	0.814 (*)	38	0.650	133	0.673 (*)	21	0.581
30	0.727	6	0.650	19	0.623	158	0.580
108	0.726	24	0.650	31	0.623	110	0.580
124	0.700	158	0.650	149	0.617	38	0.580
154	0.700	118	0.650	13	0.617	1	0.580
31	0.697	131	0.649	125	0.616	118	0.580
125	0.694	17	0.649	23	0.615	36	0.579
23	0.692	33	0.646	30	0.613	11	0.577
13	0.691	22	0.645	121	0.611	20	0.576
140	0.690	32	0.643	24	0.610	16	0.576
104	0.680	16	0.642	140	0.610	33	0.573
121	0.679	26	0.641	39	0.610	26	0.573
19	0.676	34	0.641	10	0.601	35	0.570
5	0.674	139	0.640	32	0.601	6	0.570
149	0.674	110	0.640	5	0.601	115	0.570
39	0.673	115	0.640	124	0.600	34	0.569
146	0.672	116	0.640	120	0.600	22	0.565
120	0.670	126	0.640	15	0.600	156	0.563
114	0.670	35	0.639	114	0.600	4	0.561
7	0.669	4	0.634	131	0.596	126	0.560
8	0.668	142	0.620	8	0.595	147	0.560
27	0.666	153	0.620	7	0.595	153	0.560
10	0.663	102	0.620	27	0.591	107	0.552
3	0.662	14	0.620	116	0.591	139	0.550
20	0.661	117	0.620	138	0.591	117	0.550
148	0.660	107	0.615	113	0.590	112	0.550
12	0.660	156	0.610	12	0.590	14	0.550
113	0.660	147	0.610	154	0.590	152	0.530
138	0.657	112	0.600	104	0.590	142	0.520
15	0.655	152	0.600	17	0.589	102	0.520
11	0.654	145	0.460 (*)	3	0.588	145	0.511
36	0.652	151	0.410 (*)	146	0.588	151	0.380 (*)
21	0.652			108	0.587		

Table 21: Analytical results for ammonium in precipitations sample.

Ammonium in precipitation				Ammonium in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.241				Theoretical value: 0.176			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.248				Arithmetic mean value: 0.191			
Median: 0.240				Median: 0.178			
Standard deviation 0.063				Standard deviation 0.070			
Rel. st. deviation (%) 25.533				Rel. st. deviation (%) 36.438			
Run 2:				Run 2:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.241				Arithmetic mean value: 0.183			
Median: 0.240				Median: 0.178			
Standard deviation 0.029				Standard deviation 0.030			
Rel. st. deviation (%) 11.927				Rel. st. deviation (%) 16.260			
Results in decreasing order:				Results in decreasing order:			
40	0.700 (*)	115	0.240	40	0.696 (*)	39	0.178
154	0.350	139	0.240	18	0.299	149	0.178
124	0.300	24	0.240	154	0.280	21	0.177
147	0.290	26	0.239	156	0.260	27	0.177
22	0.283	4	0.238	107	0.236	4	0.176
3	0.278	19	0.238	120	0.220	33	0.176
17	0.277	149	0.238	22	0.211	19	0.175
107	0.277	121	0.237	115	0.210	5	0.175
39	0.275	138	0.237	112	0.210	121	0.174
120	0.270	14	0.235	147	0.210	26	0.173
112	0.270	5	0.235	31	0.205	8	0.172
13	0.262	16	0.234	124	0.200	14	0.171
116	0.258	7	0.233	17	0.199	117	0.170
36	0.253	8	0.232	30	0.199	6	0.170
156	0.250	145	0.230	3	0.198	15	0.170
140	0.250	146	0.230	158	0.190	138	0.170
158	0.250	110	0.230	36	0.190	118	0.170
31	0.249	117	0.230	102	0.190	13	0.170
35	0.249	118	0.230	142	0.190	12	0.170
23	0.249	6	0.230	152	0.190	24	0.170
32	0.247	15	0.230	116	0.188	7	0.169
34	0.246	20	0.229	35	0.187	146	0.168
108	0.244	10	0.227	23	0.186	20	0.166
18	0.244	148	0.222	32	0.186	10	0.163
27	0.241	1	0.220	108	0.185	148	0.161
102	0.240	104	0.220	145	0.180	104	0.160
152	0.240	113	0.210	38	0.180	110	0.160
114	0.240	126	0.207	114	0.180	1	0.160
142	0.240	131	0.206	34	0.180	113	0.160
33	0.240	151	0.200	126	0.180	131	0.155
38	0.240	133	0.194	139	0.180	133	0.155
12	0.240	30	0.188	140	0.180	151	0.150
21	0.240	153	0.130	16	0.179	153	0.080
Ammonium in precipitation				Ammonium in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.401				Theoretical value: 0.369			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.411				Arithmetic mean value: 0.385			
Median: 0.400				Median: 0.371			
Standard deviation 0.085				Standard deviation 0.085			
Rel. st. deviation (%) 20.612				Rel. st. deviation (%) 22.105			
Run 2:				Run 2:			
Number of laboratories: 64				Number of laboratories: 65			
Arithmetic mean value: 0.399				Arithmetic mean value: 0.375			
Median: 0.400				Median: 0.371			
Standard deviation 0.034				Standard deviation 0.037			
Rel. st. deviation (%) 8.631				Rel. st. deviation (%) 9.828			
Results in decreasing order:				Results in decreasing order:			
40	0.982 (*)	104	0.400	40	0.999 (*)	21	0.371
107	0.650 (*)	39	0.399	30	0.503	24	0.370
30	0.501	121	0.399	147	0.460	110	0.370
147	0.500	5	0.399	31	0.453	27	0.369
17	0.456	33	0.399	154	0.450	4	0.368
22	0.456	26	0.399	22	0.423	108	0.367
120	0.450	4	0.397	17	0.421	26	0.367
154	0.430	118	0.395	156	0.420	5	0.367
158	0.430	108	0.395	102	0.410	149	0.365
3	0.428	14	0.394	142	0.410	118	0.363
31	0.428	19	0.391	3	0.404	121	0.362
116	0.421	7	0.390	23	0.402	14	0.360
23	0.420	115	0.390	152	0.400	104	0.360
16	0.417	117	0.390	158	0.400	117	0.360
32	0.415	24	0.390	124	0.400	19	0.358
35	0.415	8	0.389	107	0.393	15	0.358
36	0.413	10	0.387	120	0.390	131	0.358
102	0.410	146	0.382	35	0.388	8	0.358
114	0.410	148	0.382	116	0.388	13	0.356
139	0.410	138	0.381	36	0.387	10	0.355
140	0.410	13	0.381	16	0.386	146	0.355
142	0.410	15	0.380	20	0.381	7	0.353
145	0.410	6	0.380	139	0.380	138	0.352
152	0.410	34	0.376	113	0.380	1	0.350
12	0.410	156	0.373	38	0.380	6	0.350
20	0.407	1	0.370	12	0.380	34	0.350
149	0.404	126	0.365	112	0.380	148	0.344
21	0.401	131	0.357	145	0.380	126	0.330
27	0.401	113	0.350	140	0.380	18	0.327
112	0.400	18	0.339	32	0.380	115	0.320
110	0.400	151	0.320	114	0.380	151	0.300
124	0.400	133	0.318	39	0.376	133	0.295
38	0.400	153	0.290	33	0.372	153	0.270

Table 22: Analytical results for pH in precipitations samples.

pH in precipitation				pH in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 4.076				Theoretical value: 4.046			
Unit: pH-unit				Unit: pH-unit			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 4.125				Arithmetic mean value: 4.081			
Median: 4.120				Median: 4.080			
Standard deviation 0.114				Standard deviation 0.092			
Rel. st. deviation (%) 2.761				Rel. st. deviation (%) 2.261			
Run 2:				Run 2:			
Number of laboratories: 64				Number of laboratories: 63			
Arithmetic mean value: 4.118				Arithmetic mean value: 4.075			
Median: 4.120				Median: 4.080			
Standard deviation 0.068				Standard deviation 0.049			
Rel. st. deviation (%) 1.656				Rel. st. deviation (%) 1.215			
Results in decreasing order:				Results in decreasing order:			
156	4.670 (*)	118	4.120	156	4.490 (*)	24	4.080
117	4.440 (*)	120	4.120	117	4.300 (*)	8	4.080
40	4.300	24	4.110	40	4.280 (*)	23	4.080
19	4.260	107	4.110	19	4.200	13	4.070
110	4.260	121	4.110	147	4.170	18	4.070
148	4.230	16	4.110	148	4.160	21	4.070
153	4.230	3	4.108	153	4.140	39	4.070
133	4.200	124	4.100	133	4.140	107	4.070
115	4.200	15	4.100	152	4.140	139	4.070
147	4.200	139	4.100	112	4.130	5	4.070
22	4.190	21	4.100	115	4.130	4	4.067
7	4.170	39	4.100	158	4.130	22	4.060
34	4.165	116	4.100	26	4.120	126	4.060
112	4.160	4	4.096	146	4.110	33	4.050
104	4.160	5	4.090	27	4.110	121	4.050
140	4.160	23	4.080	151	4.110	6	4.050
113	4.150	33	4.080	11	4.100	12	4.050
14	4.150	32	4.080	36	4.100	16	4.050
114	4.150	30	4.080	20	4.100	116	4.050
18	4.150	6	4.080	113	4.100	32	4.040
20	4.147	149	4.070	124	4.100	149	4.040
152	4.140	10	4.060	34	4.098	145	4.030
158	4.140	12	4.060	14	4.090	154	4.030
26	4.140	142	4.050	104	4.090	10	4.030
27	4.140	102	4.050	35	4.090	1	4.030
35	4.140	145	4.050	30	4.090	142	4.010
36	4.140	1	4.020	31	4.090	7	4.010
126	4.130	138	4.010	140	4.090	102	4.010
151	4.130	108	3.990	110	4.090	138	3.990
31	4.130	131	3.970	3	4.089	108	3.950
8	4.130	154	3.970	114	4.080	131	3.940
11	4.120	17	3.930	15	4.080	17	3.940
13	4.120	38	3.720 (*)	120	4.080	38	3.690 (*)
146	4.120			118	4.080		
pH in precipitation				pH in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 4.495				Theoretical value: 4.444			
Unit: pH-unit				Unit: pH-unit			
Run 1:				Run 1:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 4.543				Arithmetic mean value: 4.500			
Median: 4.511				Median: 4.480			
Standard deviation 0.187				Standard deviation 0.197			
Rel. st. deviation (%) 4.126				Rel. st. deviation (%) 4.371			
Run 2:				Run 2:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 4.525				Arithmetic mean value: 4.481			
Median: 4.511				Median: 4.480			
Standard deviation 0.087				Standard deviation 0.072			
Rel. st. deviation (%) 1.931				Rel. st. deviation (%) 1.603			
Results in decreasing order:				Results in decreasing order:			
40	5.610 (*)	13	4.510	40	5.830 (*)	22	4.480
117	5.180 (*)	23	4.510	156	4.920 (*)	21	4.470
156	4.910	24	4.510	19	4.710	15	4.470
19	4.750	151	4.510	110	4.690	33	4.460
113	4.710	31	4.510	117	4.660	39	4.460
110	4.680	33	4.510	148	4.620	16	4.460
148	4.670	36	4.510	113	4.580	23	4.460
146	4.630	39	4.510	133	4.560	24	4.460
133	4.630	124	4.500	30	4.560	4	4.454
115	4.600	126	4.500	153	4.540	114	4.450
35	4.580	4	4.500	146	4.540	10	4.450
140	4.570	5	4.500	31	4.540	13	4.450
14	4.570	120	4.500	147	4.530	120	4.450
112	4.570	139	4.500	112	4.530	121	4.450
153	4.570	18	4.490	158	4.530	126	4.450
158	4.570	149	4.490	140	4.520	18	4.450
30	4.560	154	4.480	152	4.520	5	4.450
27	4.550	116	4.480	35	4.510	12	4.440
22	4.550	32	4.480	26	4.510	149	4.440
107	4.550	121	4.480	34	4.509	116	4.430
26	4.550	6	4.470	3	4.502	6	4.430
34	4.534	7	4.450	36	4.500	1	4.420
152	4.530	12	4.450	115	4.500	32	4.410
8	4.530	145	4.440	139	4.500	7	4.410
118	4.530	1	4.440	124	4.500	145	4.400
20	4.527	10	4.430	104	4.500	102	4.400
104	4.520	102	4.420	27	4.500	142	4.400
11	4.520	142	4.420	107	4.490	17	4.370
114	4.520	108	4.400	14	4.490	154	4.370
147	4.520	131	4.380	118	4.480	108	4.360
21	4.520	138	4.380	151	4.480	131	4.350
16	4.520	17	4.370	8	4.480	138	4.340
15	4.520	38	4.020 (*)	11	4.480	38	3.990 (*)
3	4.511			20	4.480		

Table 23: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH		Strong acid calculated from pH	
Sample no.: 1		Sample no.: 2	
Theoretical value:	84.000	Theoretical value:	90.000
Unit: µeq		Unit: µeq	
Run 1:		Run 1:	
Number of laboratories:	67	Number of laboratories:	67
Arithmetic mean value:	77.386	Arithmetic mean value:	84.772
Median:	75.858	Median:	83.176
Standard deviation	20.320	Standard deviation	19.633
Rel. st. deviation (%)	26.258	Rel. st. deviation (%)	23.159
Run 2:		Run 2:	
Number of laboratories:	64	Number of laboratories:	65
Arithmetic mean value:	77.135	Arithmetic mean value:	83.741
Median:	75.858	Median:	83.176
Standard deviation	12.357	Standard deviation	11.433
Rel. st. deviation (%)	16.019	Rel. st. deviation (%)	13.653
Results in decreasing order:		Results in decreasing order:	
38 190.546 (*)	13 75.858	38 204.174 (*)	15 83.176
17 117.490	118 75.858	17 114.815	120 83.176
154 107.152	8 74.195	131 114.815	8 82.272
131 107.152	151 74.131	108 112.202	3 81.470
108 102.329	126 74.131	138 102.329	110 81.283
138 97.724	31 74.131	142 97.724	14 81.283
1 95.499	26 72.444	7 97.724	140 81.283
142 89.125	152 72.444	102 97.724	31 81.283
145 89.125	158 72.444	1 93.325	104 81.283
102 89.125	27 72.444	154 93.325	35 81.283
10 87.096	36 72.444	10 93.325	30 81.283
12 87.096	35 72.444	145 93.325	34 79.799
149 85.114	20 71.285	149 91.201	11 79.433
32 83.176	14 70.795	32 91.201	113 79.433
33 83.176	18 70.795	6 89.125	20 79.433
6 83.176	113 70.795	12 89.125	124 79.433
23 83.176	114 70.795	33 89.125	36 79.433
30 83.176	104 69.183	121 89.125	151 77.625
5 81.283	140 69.183	116 89.125	27 77.625
4 80.168	112 69.183	16 89.125	146 77.625
15 79.433	34 68.391	126 87.096	26 75.858
139 79.433	7 67.608	22 87.096	158 74.131
39 79.433	22 64.565	4 85.704	115 74.131
116 79.433	133 63.096	18 85.114	112 74.131
124 79.433	115 63.096	5 85.114	152 72.444
21 79.433	147 63.096	139 85.114	153 72.444
3 77.983	153 58.884	21 85.114	133 72.444
107 77.625	148 58.884	13 85.114	148 69.183
16 77.625	110 54.954	39 85.114	147 67.608
121 77.625	19 54.954	107 85.114	19 63.096
24 77.625	40 50.119	24 83.176	40 52.481
146 75.858	117 36.308 (*)	114 83.176	117 50.119
120 75.858	156 21.380 (*)	118 83.176	156 32.359 (*)
11 75.858		23 83.176	
Strong acid calculated from pH		Strong acid calculated from pH	
Sample no.: 3		Sample no.: 4	
Theoretical value:	32.000	Theoretical value:	36.000
Unit: µeq		Unit: µeq	
Run 1:		Run 1:	
Number of laboratories:	67	Number of laboratories:	67
Arithmetic mean value:	30.596	Arithmetic mean value:	33.692
Median:	30.832	Median:	33.113
Standard deviation	10.611	Standard deviation	10.973
Rel. st. deviation (%)	34.682	Rel. st. deviation (%)	32.570
Run 2:		Run 2:	
Number of laboratories:	64	Number of laboratories:	65
Arithmetic mean value:	30.396	Arithmetic mean value:	33.131
Median:	30.867	Median:	33.113
Standard deviation	5.418	Standard deviation	5.834
Rel. st. deviation (%)	17.823	Rel. st. deviation (%)	17.608
Results in decreasing order:		Results in decreasing order:	
38 95.499 (*)	15 30.200	38 102.329 (*)	151 33.113
17 42.658	104 30.200	138 45.709	20 33.113
138 41.687	16 30.200	131 44.668	118 33.113
131 41.687	21 30.200	108 43.652	8 32.933
108 39.811	11 30.200	154 42.658	14 32.359
102 38.019	147 30.200	17 42.658	107 32.359
142 38.019	114 30.200	145 39.811	115 31.623
10 37.154	20 29.717	102 39.811	27 31.623
1 36.308	152 29.512	142 39.811	104 31.623
145 36.308	118 29.512	7 38.905	36 31.623
7 35.481	8 29.478	32 38.905	124 31.623
12 35.481	34 29.242	1 38.019	139 31.623
6 33.884	107 28.184	116 37.154	3 31.477
154 33.113	26 28.184	6 37.154	34 30.974
116 33.113	27 28.184	149 36.308	26 30.903
121 33.113	22 28.184	12 36.308	35 30.903
32 33.113	30 27.542	13 35.481	152 30.200
149 32.359	153 26.915	18 35.481	140 30.200
18 32.359	158 26.915	10 35.481	158 29.512
4 31.623	112 26.915	5 35.481	112 29.512
5 31.623	140 26.915	120 35.481	147 29.512
139 31.623	14 26.915	114 35.481	31 28.840
124 31.623	35 26.303	126 35.481	146 28.840
126 31.623	115 25.119	121 35.481	153 28.840
120 31.623	133 23.442	4 35.156	30 27.542
39 30.903	146 23.442	23 34.674	133 27.542
151 30.903	148 21.380	24 34.674	113 26.303
13 30.903	110 20.893	33 34.674	148 23.988
23 30.903	113 19.498	39 34.674	117 21.878
24 30.903	19 17.783	16 34.674	110 20.417
31 30.903	156 12.303	15 33.884	19 19.498
33 30.903	117 6.607 (*)	21 33.884	156 12.023
36 30.903	40 2.455 (*)	11 33.113	40 1.479 (*)
3 30.832		22 33.113	

Table 24: Analytical results for chloride in precipitations samples.

Chloride in precipitation				Chloride in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.185				Theoretical value: 0.232			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 66			
Arithmetic mean value: 0.195				Arithmetic mean value: 0.229			
Median: 0.180				Median: 0.227			
Standard deviation 0.137				Standard deviation 0.069			
Rel. st. deviation (%) 70.438				Rel. st. deviation (%) 30.108			
Run 2:				Run 2:			
Number of laboratories: 63				Number of laboratories: 62			
Arithmetic mean value: 0.179				Arithmetic mean value: 0.227			
Median: 0.180				Median: 0.227			
Standard deviation 0.053				Standard deviation 0.035			
Rel. st. deviation (%) 29.905				Rel. st. deviation (%) 15.355			
Results in decreasing order:				Results in decreasing order:			
156	1.190 (*)	6	0.180	156	0.576 (*)	138	0.226
17	0.334	140	0.180	17	0.390 (*)	34	0.225
154	0.320	146	0.180	133	0.328	146	0.224
133	0.286	7	0.179	10	0.301	11	0.223
10	0.270	16	0.178	120	0.300	35	0.223
115	0.230	34	0.178	154	0.300	23	0.223
120	0.230	32	0.175	113	0.280	158	0.220
113	0.230	3	0.173	39	0.274	38	0.220
139	0.220	4	0.173	108	0.260	114	0.220
39	0.212	138	0.170	139	0.260	6	0.220
24	0.210	38	0.170	115	0.260	16	0.218
110	0.210	13	0.170	40	0.259	4	0.216
145	0.206	1	0.170	117	0.250	3	0.215
40	0.205	158	0.170	110	0.250	32	0.215
124	0.200	31	0.164	140	0.250	22	0.215
12	0.200	20	0.164	30	0.243	126	0.210
117	0.200	5	0.163	24	0.240	1	0.210
153	0.200	131	0.163	153	0.240	20	0.209
152	0.200	107	0.161	21	0.239	5	0.206
23	0.198	126	0.160	131	0.237	31	0.202
30	0.197	151	0.160	107	0.235	145	0.201
33	0.193	11	0.158	26	0.235	124	0.200
149	0.188	104	0.150	15	0.235	152	0.200
21	0.185	108	0.140	121	0.232	148	0.200
8	0.184	14	0.130	8	0.231	118	0.190
36	0.183	147	0.130	13	0.231	116	0.185
35	0.183	116	0.130	36	0.231	104	0.180
26	0.183	22	0.102	12	0.230	151	0.180
27	0.182	18	0.075	149	0.230	147	0.170
118	0.180	102	0.040	125	0.228	14	0.170
15	0.180	142	0.040	27	0.228	18	0.100
114	0.180	148	0.010	7	0.227	142	0.020 (*)
		19	<0.300	33	0.227	102	0.020 (*)
		112	<0.370			19	<0.300
		121	<0.100			112	<0.370
		125	<0.200				
Chloride in precipitation				Chloride in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.533				Theoretical value: 0.556			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 68				Number of laboratories: 68			
Arithmetic mean value: 0.540				Arithmetic mean value: 0.558			
Median: 0.510				Median: 0.532			
Standard deviation 0.203				Standard deviation 0.182			
Rel. st. deviation (%) 37.535				Rel. st. deviation (%) 32.690			
Run 2:				Run 2:			
Number of laboratories: 67				Number of laboratories: 67			
Arithmetic mean value: 0.517				Arithmetic mean value: 0.539			
Median: 0.510				Median: 0.532			
Standard deviation 0.069				Standard deviation 0.088			
Rel. st. deviation (%) 13.414				Rel. st. deviation (%) 16.327			
Results in decreasing order:				Results in decreasing order:			
156	2.090 (*)	158	0.510	156	1.860 (*)	20	0.532
17	0.847	142	0.510	17	0.870	146	0.531
133	0.714	6	0.510	133	0.864	153	0.530
10	0.641	153	0.510	10	0.845	6	0.530
40	0.610	102	0.510	24	0.630	158	0.530
39	0.592	34	0.509	22	0.629	138	0.529
108	0.580	36	0.505	39	0.627	36	0.526
139	0.580	131	0.502	140	0.590	107	0.521
22	0.570	20	0.502	115	0.580	116	0.521
154	0.570	1	0.500	18	0.577	1	0.520
24	0.560	124	0.500	40	0.575	38	0.520
115	0.550	38	0.500	15	0.570	4	0.517
26	0.546	117	0.500	139	0.570	13	0.515
15	0.545	145	0.499	26	0.567	32	0.515
12	0.540	30	0.498	21	0.565	16	0.512
8	0.536	138	0.497	125	0.565	118	0.510
125	0.535	32	0.497	35	0.564	112	0.510
149	0.532	13	0.493	12	0.560	145	0.506
27	0.531	16	0.490	113	0.560	19	0.502
140	0.530	120	0.490	8	0.557	31	0.501
110	0.530	4	0.487	7	0.556	152	0.500
7	0.530	19	0.484	27	0.553	124	0.500
113	0.530	116	0.484	149	0.550	108	0.490
23	0.529	118	0.480	110	0.550	121	0.485
35	0.529	31	0.476	114	0.550	30	0.479
21	0.529	112	0.470	33	0.550	14	0.460
5	0.527	18	0.451	34	0.544	117	0.450
33	0.524	121	0.450	5	0.542	102	0.440
146	0.521	14	0.440	126	0.540	142	0.440
126	0.520	147	0.430	154	0.540	147	0.430
114	0.520	148	0.430	3	0.539	120	0.430
11	0.517	152	0.400	131	0.538	104	0.420
3	0.517	104	0.380	23	0.534	148	0.390
107	0.511	151	0.310	11	0.533	151	0.320

Table 25: Analytical results for sodium in precipitations samples.

Sodium in precipitation				Sodium in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.271				Theoretical value: 0.322			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.266				Arithmetic mean value: 0.319			
Median: 0.269				Median: 0.319			
Standard deviation 0.042				Standard deviation 0.041			
Rel. st. deviation (%) 15.660				Rel. st. deviation (%) 12.865			
Run 2:				Run 2:			
Number of laboratories: 62				Number of laboratories: 62			
Arithmetic mean value: 0.260				Arithmetic mean value: 0.318			
Median: 0.269				Median: 0.319			
Standard deviation 0.025				Standard deviation 0.028			
Rel. st. deviation (%) 9.671				Rel. st. deviation (%) 8.890			
Results in decreasing order:				Results in decreasing order:			
148	0.470 (*)	138	0.269	148	0.491 (*)	19	0.318
156	0.390 (*)	7	0.268	35	0.418 (*)	107	0.316
149	0.354 (*)	107	0.265	140	0.400	8	0.315
4	0.314	31	0.263	156	0.380	38	0.313
40	0.304	38	0.261	32	0.375	124	0.310
147	0.300	8	0.261	34	0.367	16	0.310
32	0.298	16	0.260	4	0.366	21	0.308
133	0.295	124	0.260	14	0.364	39	0.308
14	0.285	104	0.260	108	0.346	30	0.306
146	0.284	39	0.254	121	0.340	133	0.305
1	0.280	21	0.250	118	0.340	20	0.301
112	0.280	153	0.250	126	0.340	115	0.300
126	0.280	115	0.250	158	0.340	13	0.300
151	0.280	11	0.247	15	0.336	102	0.300
15	0.280	10	0.246	146	0.336	104	0.300
140	0.280	152	0.245	40	0.335	142	0.300
6	0.280	30	0.242	149	0.335	120	0.299
121	0.280	116	0.242	145	0.334	11	0.299
158	0.280	117	0.240	3	0.333	22	0.296
34	0.276	24	0.240	36	0.332	152	0.295
26	0.276	102	0.240	31	0.331	10	0.294
5	0.275	142	0.240	33	0.331	154	0.290
145	0.274	110	0.240	138	0.331	153	0.290
19	0.274	22	0.238	151	0.330	110	0.290
33	0.272	120	0.234	112	0.330	24	0.280
13	0.271	154	0.230	6	0.330	117	0.280
36	0.271	113	0.220	26	0.326	116	0.274
114	0.270	131	0.217	5	0.325	23	0.274
3	0.270	35	0.214	114	0.320	113	0.270
20	0.270	23	0.204	12	0.320	131	0.270
12	0.270	108	0.197	1	0.320	17	0.257
118	0.270	17	0.195	7	0.319	139	0.222 (*)
27	0.269	139	0.181 (*)	27	0.319	147	0.210 (*)
Sodium in precipitation				Sodium in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.668				Theoretical value: 0.640			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.655				Arithmetic mean value: 0.622			
Median: 0.655				Median: 0.624			
Standard deviation 0.056				Standard deviation 0.063			
Rel. st. deviation (%) 8.599				Rel. st. deviation (%) 10.071			
Run 2:				Run 2:			
Number of laboratories: 62				Number of laboratories: 64			
Arithmetic mean value: 0.654				Arithmetic mean value: 0.620			
Median: 0.655				Median: 0.624			
Standard deviation 0.036				Standard deviation 0.048			
Rel. st. deviation (%) 5.548				Rel. st. deviation (%) 7.745			
Results in decreasing order:				Results in decreasing order:			
148	0.879 (*)	107	0.655	148	0.897 (*)	107	0.624
30	0.801 (*)	108	0.654	30	0.742	38	0.623
4	0.722	33	0.654	4	0.700	8	0.623
32	0.720	8	0.652	156	0.690	118	0.620
14	0.715	27	0.651	14	0.689	151	0.610
156	0.710	118	0.650	133	0.685	104	0.610
147	0.710	152	0.649	35	0.680	11	0.607
35	0.709	3	0.645	3	0.675	20	0.605
15	0.701	39	0.642	15	0.674	138	0.604
133	0.700	117	0.640	121	0.670	39	0.603
158	0.700	151	0.640	126	0.670	102	0.600
149	0.692	23	0.634	147	0.670	110	0.600
126	0.690	11	0.632	158	0.670	153	0.600
34	0.684	104	0.630	145	0.668	117	0.600
140	0.680	153	0.630	32	0.665	142	0.600
121	0.680	24	0.630	149	0.659	13	0.596
145	0.676	115	0.630	120	0.659	24	0.590
146	0.675	138	0.629	140	0.650	152	0.588
36	0.675	20	0.627	36	0.649	10	0.587
26	0.674	102	0.620	26	0.643	131	0.587
5	0.673	142	0.620	5	0.642	23	0.585
112	0.670	13	0.618	112	0.640	31	0.583
110	0.670	10	0.618	1	0.640	40	0.576
1	0.670	40	0.616	21	0.640	124	0.570
31	0.668	131	0.614	114	0.640	115	0.560
19	0.667	120	0.606	146	0.639	113	0.540
21	0.664	124	0.590	7	0.631	108	0.537
6	0.660	116	0.589	6	0.630	34	0.535
16	0.660	22	0.585	19	0.630	17	0.533
114	0.660	17	0.577	16	0.630	22	0.522
12	0.660	113	0.560	12	0.630	154	0.520
7	0.658	154	0.466 (*)	33	0.625	116	0.511
38	0.656	139	0.466 (*)	27	0.624	139	0.439 (*)

Table 26: Analytical results for magnesium in precipitations samples.

Magnesium in precipitation				Magnesium in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.087				Theoretical value: 0.099			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 65			
Arithmetic mean value: 0.088				Arithmetic mean value: 0.099			
Median: 0.086				Median: 0.099			
Standard deviation 0.028				Standard deviation 0.032			
Rel. st. deviation (%) 31.406				Rel. st. deviation (%) 32.716			
Run 2:				Run 2:			
Number of laboratories: 61				Number of laboratories: 62			
Arithmetic mean value: 0.084				Arithmetic mean value: 0.095			
Median: 0.086				Median: 0.099			
Standard deviation 0.012				Standard deviation 0.012			
Rel. st. deviation (%) 14.036				Rel. st. deviation (%) 12.781			
Results in decreasing order:				Results in decreasing order:			
154	0.230 (*)	146	0.086	154	0.290 (*)	26	0.099
147	0.210 (*)	40	0.086	147	0.220 (*)	131	0.098
113	0.120	5	0.085	113	0.120	14	0.098
126	0.100	131	0.085	30	0.115	8	0.098
153	0.100	3	0.085	102	0.110	19	0.098
102	0.100	33	0.084	126	0.110	10	0.098
145	0.100	121	0.084	153	0.110	146	0.098
142	0.100	4	0.084	142	0.110	120	0.096
138	0.097	14	0.083	138	0.108	121	0.095
149	0.095	16	0.083	145	0.108	33	0.095
107	0.093	151	0.083	107	0.105	39	0.095
1	0.090	120	0.083	17	0.102	151	0.095
104	0.090	23	0.081	139	0.101	16	0.094
140	0.090	11	0.081	140	0.100	36	0.094
15	0.090	31	0.081	149	0.100	23	0.094
17	0.090	6	0.080	21	0.100	11	0.094
124	0.090	7	0.080	5	0.100	1	0.090
118	0.090	30	0.080	156	0.100	6	0.090
117	0.090	115	0.080	34	0.100	7	0.090
112	0.090	114	0.080	104	0.100	24	0.090
139	0.089	156	0.080	112	0.100	115	0.090
12	0.088	158	0.080	114	0.100	158	0.090
26	0.088	20	0.078	117	0.100	38	0.086
10	0.088	38	0.074	13	0.100	22	0.085
13	0.088	22	0.073	118	0.100	20	0.079
34	0.088	24	0.070	15	0.100	40	0.077
152	0.087	148	0.061	124	0.100	148	0.072
27	0.087	110	0.060	27	0.099	110	0.070
36	0.087	39	0.054	3	0.099	116	0.068
21	0.087	133	0.051	4	0.099	133	0.058
19	0.087	35	0.048	31	0.099	35	0.051
8	0.087	116	0.028 (*)	12	0.099	108	0.019 (*)
		108	<0.001	152	0.099		
Magnesium in precipitation				Magnesium in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.124				Theoretical value: 0.136			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.127				Arithmetic mean value: 0.140			
Median: 0.123				Median: 0.136			
Standard deviation 0.045				Standard deviation 0.047			
Rel. st. deviation (%) 35.665				Rel. st. deviation (%) 33.559			
Run 2:				Run 2:			
Number of laboratories: 63				Number of laboratories: 63			
Arithmetic mean value: 0.119				Arithmetic mean value: 0.132			
Median: 0.123				Median: 0.136			
Standard deviation 0.017				Standard deviation 0.018			
Rel. st. deviation (%) 14.574				Rel. st. deviation (%) 13.766			
Results in decreasing order:				Results in decreasing order:			
147	0.400 (*)	8	0.123	147	0.410 (*)	27	0.136
154	0.310 (*)	5	0.122	154	0.350 (*)	5	0.135
140	0.150	31	0.122	30	0.199	10	0.135
30	0.142	7	0.121	153	0.160	19	0.135
107	0.142	118	0.121	113	0.150	14	0.134
113	0.140	121	0.120	107	0.150	131	0.134
153	0.140	15	0.120	138	0.149	15	0.134
126	0.140	6	0.120	17	0.147	152	0.134
138	0.138	3	0.120	31	0.144	120	0.133
17	0.133	1	0.120	145	0.143	7	0.132
104	0.130	39	0.120	104	0.140	39	0.131
102	0.130	14	0.119	102	0.140	158	0.130
112	0.130	33	0.118	112	0.140	121	0.130
117	0.130	23	0.117	117	0.140	124	0.130
142	0.130	120	0.117	126	0.140	6	0.130
24	0.130	151	0.116	140	0.140	1	0.130
156	0.130	16	0.115	142	0.140	151	0.129
145	0.128	11	0.115	23	0.140	33	0.129
12	0.128	115	0.112	24	0.140	11	0.128
146	0.126	114	0.110	12	0.140	16	0.126
149	0.126	22	0.110	156	0.140	115	0.123
139	0.126	38	0.110	26	0.138	38	0.123
131	0.125	124	0.110	3	0.138	148	0.123
21	0.124	158	0.110	4	0.138	20	0.121
19	0.124	20	0.107	139	0.138	114	0.120
152	0.124	116	0.103	149	0.138	22	0.120
4	0.124	40	0.097	146	0.137	116	0.116
13	0.124	148	0.095	36	0.137	40	0.107
34	0.124	110	0.090	21	0.137	110	0.100
27	0.124	35	0.078	118	0.137	35	0.082
10	0.123	133	0.072	8	0.136	133	0.078
26	0.123	108	0.037	34	0.136	108	0.063
36	0.123			13	0.136		

Table 27: Analytical results for calcium in precipitations samples.

Calcium in precipitation				Calcium in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.199				Theoretical value: 0.238			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.213				Arithmetic mean value: 0.246			
Median: 0.203				Median: 0.240			
Standard deviation 0.111				Standard deviation 0.110			
Rel. st. deviation (%) 52.313				Rel. st. deviation (%) 44.837			
Run 2:				Run 2:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.200				Arithmetic mean value: 0.234			
Median: 0.202				Median: 0.240			
Standard deviation 0.050				Standard deviation 0.048			
Rel. st. deviation (%) 24.870				Rel. st. deviation (%) 20.684			
Results in decreasing order:				Results in decreasing order:			
147	1.010 (*)	27	0.202	147	1.040 (*)	16	0.240
126	0.370	6	0.200	39	0.377	4	0.240
39	0.326	1	0.200	156	0.340	27	0.239
156	0.300	15	0.200	108	0.317	8	0.238
40	0.287	120	0.198	126	0.310	146	0.236
113	0.260	21	0.195	145	0.295	13	0.235
145	0.257	12	0.190	153	0.290	10	0.233
114	0.240	102	0.190	40	0.279	38	0.232
149	0.237	104	0.190	31	0.275	1	0.230
31	0.232	115	0.190	30	0.261	140	0.230
153	0.230	151	0.190	34	0.260	124	0.230
140	0.230	38	0.190	112	0.260	24	0.230
19	0.226	142	0.190	114	0.260	139	0.224
152	0.220	139	0.189	107	0.253	23	0.222
118	0.220	23	0.188	3	0.251	33	0.221
112	0.220	33	0.188	118	0.250	115	0.220
107	0.216	20	0.186	152	0.250	151	0.220
36	0.216	133	0.185	121	0.250	12	0.220
138	0.214	30	0.182	113	0.250	104	0.220
3	0.213	11	0.182	138	0.250	158	0.220
146	0.213	7	0.182	117	0.250	131	0.217
10	0.211	158	0.180	36	0.246	11	0.216
121	0.210	24	0.180	26	0.245	7	0.215
34	0.210	108	0.177	19	0.245	110	0.210
4	0.210	131	0.175	14	0.243	20	0.191
117	0.210	124	0.170	21	0.242	32	0.180
110	0.210	17	0.128	5	0.242	35	0.177
26	0.208	22	0.128	149	0.242	17	0.175
8	0.208	32	0.125	15	0.240	133	0.161
16	0.208	154	0.120	120	0.240	22	0.155
13	0.205	35	0.109	102	0.240	154	0.130
5	0.204	116	0.055	6	0.240	148	0.077
14	0.203	148	0.048	142	0.240	116	0.069
Calcium in precipitation				Calcium in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.276				Theoretical value: 0.284			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories: 66				Number of laboratories: 66			
Arithmetic mean value: 0.297				Arithmetic mean value: 0.306			
Median: 0.278				Median: 0.286			
Standard deviation 0.209				Standard deviation 0.195			
Rel. st. deviation (%) 70.333				Rel. st. deviation (%) 63.848			
Run 2:				Run 2:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.272				Arithmetic mean value: 0.283			
Median: 0.277				Median: 0.286			
Standard deviation 0.056				Standard deviation 0.054			
Rel. st. deviation (%) 20.486				Rel. st. deviation (%) 19.089			
Results in decreasing order:				Results in decreasing order:			
147	1.910 (*)	8	0.277	147	1.810 (*)	146	0.286
35	0.435	16	0.277	39	0.451	16	0.285
39	0.427	3	0.276	35	0.412	149	0.285
153	0.350	146	0.276	153	0.360	32	0.283
107	0.343	27	0.275	156	0.350	27	0.283
108	0.341	14	0.275	140	0.350	8	0.283
156	0.340	13	0.274	30	0.347	10	0.281
38	0.333	6	0.270	31	0.335	142	0.280
30	0.321	4	0.270	107	0.322	4	0.280
140	0.310	120	0.268	145	0.319	24	0.280
126	0.310	7	0.266	126	0.310	102	0.280
19	0.305	131	0.264	114	0.310	13	0.279
31	0.300	12	0.260	138	0.305	131	0.276
138	0.295	24	0.260	34	0.305	133	0.274
149	0.290	139	0.260	120	0.304	38	0.273
34	0.290	151	0.260	19	0.303	7	0.272
121	0.290	33	0.252	23	0.302	12	0.270
118	0.290	23	0.251	36	0.301	151	0.270
113	0.290	32	0.250	121	0.300	104	0.270
112	0.290	104	0.250	118	0.300	158	0.270
152	0.290	11	0.250	112	0.300	139	0.266
145	0.290	115	0.250	108	0.297	33	0.263
36	0.285	158	0.250	3	0.295	11	0.261
10	0.284	110	0.240	26	0.295	110	0.260
26	0.281	124	0.240	1	0.290	115	0.260
1	0.280	20	0.238	6	0.290	20	0.248
102	0.280	40	0.206	21	0.290	124	0.240
114	0.280	22	0.202	113	0.290	40	0.216
5	0.280	17	0.196	15	0.290	17	0.205
117	0.280	133	0.176	117	0.290	22	0.190
142	0.280	154	0.150	152	0.290	154	0.140
15	0.280	116	0.093	5	0.289	148	0.124
21	0.279	148	0.085	14	0.287	116	0.082

Table 28: Analytical results for potassium in precipitations samples.

Potassium in precipitation		Potassium in precipitation	
Sample no.: 1		Sample no.: 2	
Theoretical value:	0.306	Theoretical value:	0.265
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	65	Number of laboratories:	64
Arithmetic mean value:	0.297	Arithmetic mean value:	0.258
Median:	0.297	Median:	0.257
Standard deviation	0.039	Standard deviation	0.040
Rel. st. deviation (%)	13.119	Rel. st. deviation (%)	15.427
Run 2:		Run 2:	
Number of laboratories:	61	Number of laboratories:	61
Arithmetic mean value:	0.293	Arithmetic mean value:	0.256
Median:	0.296	Median:	0.256
Standard deviation	0.025	Standard deviation	0.027
Rel. st. deviation (%)	8.600	Rel. st. deviation (%)	10.543
Results in decreasing order:		Results in decreasing order:	
149 0.458 (*) 39 0.296		153 0.410 (*) 34 0.256	
153 0.400 (*) 13 0.296		149 0.390 (*) 39 0.256	
148 0.377 (*) 21 0.295		148 0.331 21 0.256	
14 0.360 40 0.295		156 0.320 152 0.253	
112 0.350 146 0.293		112 0.320 7 0.252	
156 0.350 23 0.292		14 0.307 33 0.252	
11 0.345 113 0.290		108 0.304 145 0.251	
10 0.324 115 0.290		110 0.280 146 0.251	
117 0.320 33 0.290		10 0.280 32 0.250	
19 0.317 16 0.290		138 0.280 158 0.250	
5 0.314 151 0.290		5 0.270 113 0.250	
26 0.313 158 0.290		117 0.270 114 0.250	
126 0.310 32 0.289		26 0.270 121 0.250	
118 0.310 3 0.289		118 0.270 124 0.250	
15 0.310 36 0.286		31 0.270 104 0.250	
138 0.307 31 0.284		3 0.269 147 0.250	
1 0.304 116 0.283		19 0.269 13 0.248	
8 0.302 104 0.280		40 0.269 120 0.247	
34 0.301 124 0.280		23 0.265 16 0.244	
121 0.300 145 0.278		8 0.262 116 0.244	
4 0.300 140 0.270		1 0.260 20 0.241	
6 0.300 38 0.262		4 0.260 30 0.237	
114 0.300 30 0.258		6 0.260 131 0.227	
12 0.300 131 0.256		142 0.260 35 0.219	
147 0.300 133 0.255		12 0.260 38 0.217	
110 0.300 35 0.252		151 0.260 139 0.215	
107 0.300 154 0.250		15 0.260 22 0.211	
142 0.300 120 0.249		102 0.260 115 0.210	
102 0.300 108 0.247		126 0.260 140 0.210	
27 0.300 139 0.243		27 0.258 133 0.205	
152 0.298 22 0.239		36 0.258 154 0.180	
20 0.297 17 0.174 (*)		107 0.258 17 0.133 (*)	
7 0.297		11 <0.300	
Potassium in precipitation		Potassium in precipitation	
Sample no.: 3		Sample no.: 4	
Theoretical value:	0.183	Theoretical value:	0.143
Unit: µg/l		Unit: µg/l	
Run 1:		Run 1:	
Number of laboratories:	64	Number of laboratories:	64
Arithmetic mean value:	0.185	Arithmetic mean value:	0.144
Median:	0.180	Median:	0.139
Standard deviation	0.041	Standard deviation	0.040
Rel. st. deviation (%)	22.227	Rel. st. deviation (%)	27.549
Run 2:		Run 2:	
Number of laboratories:	59	Number of laboratories:	59
Arithmetic mean value:	0.179	Arithmetic mean value:	0.139
Median:	0.180	Median:	0.138
Standard deviation	0.022	Standard deviation	0.024
Rel. st. deviation (%)	12.228	Rel. st. deviation (%)	16.945
Results in decreasing order:		Results in decreasing order:	
153 0.320 (*) 110 0.180		120 0.269 (*) 13 0.138	
30 0.308 (*) 114 0.180		149 0.263 (*) 27 0.138	
149 0.306 (*) 118 0.180		156 0.260 (*) 39 0.138	
156 0.300 (*) 8 0.179		153 0.230 (*) 40 0.138	
108 0.244 13 0.179		148 0.210 107 0.137	
148 0.231 31 0.179		112 0.190 8 0.137	
112 0.230 20 0.178		30 0.181 23 0.137	
121 0.220 7 0.178		121 0.180 5 0.137	
151 0.210 27 0.177		151 0.180 33 0.136	
22 0.201 39 0.177		108 0.179 7 0.133	
115 0.200 23 0.177		31 0.166 114 0.130	
14 0.193 3 0.176		4 0.160 142 0.130	
19 0.192 33 0.175		14 0.158 140 0.130	
158 0.190 116 0.174		34 0.156 124 0.130	
117 0.190 107 0.173		19 0.152 126 0.130	
12 0.190 6 0.170		12 0.150 16 0.130	
10 0.189 124 0.170		117 0.150 147 0.130	
26 0.187 36 0.169		146 0.147 6 0.130	
146 0.186 152 0.167		3 0.146 102 0.130	
34 0.185 145 0.166		138 0.146 110 0.130	
133 0.185 16 0.166		10 0.145 36 0.127	
138 0.184 120 0.165		26 0.144 152 0.126	
5 0.184 131 0.162		145 0.143 116 0.124	
32 0.181 102 0.160		20 0.141 22 0.120	
1 0.180 142 0.160		118 0.140 131 0.116	
126 0.180 140 0.160		1 0.140 139 0.115	
4 0.180 139 0.150		113 0.140 35 0.114	
147 0.180 35 0.147		104 0.140 133 0.100	
15 0.180 38 0.133		21 0.140 38 0.094	
21 0.180 113 0.130		15 0.140 154 0.080	
40 0.180 154 0.110		158 0.140 115 0.070	
104 0.180 17 0.060 (*)		32 0.139 17 0.020 (*)	
11 <0.300		11 <0.300	

Table 29: Analytical results for conductivity in precipitations samples.

Conductivity in precipitation				Conductivity in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 43.500				Theoretical value: 45.900			
Unit: µS/cm				Unit: µS/cm			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 63			
Arithmetic mean value: 40.694				Arithmetic mean value: 43.095			
Median: 42.000				Median: 44.800			
Standard deviation 6.135				Standard deviation 6.381			
Rel. st. deviation (%) 15.075				Rel. st. deviation (%) 14.806			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 60			
Arithmetic mean value: 41.789				Arithmetic mean value: 44.237			
Median: 42.200				Median: 44.950			
Standard deviation 2.911				Standard deviation 2.923			
Rel. st. deviation (%) 6.965				Rel. st. deviation (%) 6.607			
Results in decreasing order:				Results in decreasing order:			
139	50.000	117	42.000	139	53.000	18	44.800
149	49.880	23	41.900	12	48.400	4	44.700
108	45.700	158	41.700	108	47.800	23	44.600
12	45.600	114	41.700	5	47.000	13	44.233
5	44.500	126	41.400	140	46.700	34	44.200
21	44.060	32	41.400	21	46.520	16	44.000
142	43.700	121	41.300	20	46.400	151	44.000
102	43.700	133	41.300	38	46.300	8	43.900
115	43.600	8	41.300	115	46.000	36	43.900
20	43.530	145	41.200	14	46.000	133	43.900
10	43.500	18	41.000	10	45.850	32	43.700
31	43.400	124	41.000	15	45.700	110	43.700
7	43.200	118	41.000	31	45.700	30	43.500
15	43.200	154	40.900	11	45.500	145	43.400
14	43.000	16	40.900	114	45.400	158	43.300
151	42.900	120	40.800	22	45.400	6	43.100
107	42.800	34	40.500	27	45.300	118	43.000
3	42.800	17	40.500	3	45.300	17	43.000
38	42.800	110	40.500	102	45.300	156	42.600
1	42.800	6	39.800	142	45.300	124	42.000
27	42.800	146	39.700	121	45.200	147	42.000
112	42.700	152	38.600	126	45.200	152	41.400
33	42.700	104	38.300	104	45.200	153	41.290
22	42.600	147	37.000	1	45.100	146	41.100
19	42.500	156	36.800	112	45.100	113	38.700
36	42.400	113	34.800	35	45.100	7	38.000
13	42.400	153	34.470	19	45.100	116	37.410
140	42.300	116	31.910	33	45.100	148	31.000
11	42.200	148	26.200 (*)	107	45.100	40	29.000 (*)
35	42.200	40	26.000 (*)	117	45.000	149	27.406 (*)
4	42.200	138	4.150 (*)	120	44.900	138	4.400 (*)
30	42.000			154	44.800		
Conductivity in precipitation				Conductivity in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 25.500				Theoretical value: 26.800			
Unit: µS/cm				Unit: µS/cm			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 63			
Arithmetic mean value: 24.686				Arithmetic mean value: 24.905			
Median: 25.200				Median: 26.030			
Standard deviation 3.490				Standard deviation 4.900			
Rel. st. deviation (%) 14.136				Rel. st. deviation (%) 19.676			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 61			
Arithmetic mean value: 25.062				Arithmetic mean value: 25.660			
Median: 25.250				Median: 26.100			
Standard deviation 1.370				Standard deviation 2.547			
Rel. st. deviation (%) 5.465				Rel. st. deviation (%) 9.928			
Results in decreasing order:				Results in decreasing order:			
139	33.000 (*)	112	25.200	139	34.000	117	26.000
149	28.672	115	25.100	3	28.100	23	26.000
34	26.700	120	25.100	108	27.800	115	26.000
108	26.700	23	25.100	5	27.700	32	25.900
12	26.600	147	25.000	12	27.700	36	25.880
31	26.500	118	25.000	104	27.200	6	25.800
153	26.480	117	25.000	15	27.200	107	25.800
5	26.400	32	24.900	21	27.130	8	25.800
104	26.300	6	24.900	140	27.000	147	25.600
3	26.270	133	24.800	34	27.000	35	25.600
121	26.200	35	24.800	20	26.960	110	25.600
30	26.200	110	24.800	22	26.900	158	25.600
15	26.100	13	24.700	121	26.900	16	25.500
14	26.000	8	24.700	10	26.720	102	25.300
21	25.880	158	24.600	27	26.700	142	25.300
20	25.860	107	24.500	4	26.700	145	25.100
11	25.800	152	24.400	31	26.600	7	25.100
10	25.730	16	24.300	126	26.600	14	25.000
27	25.700	145	24.200	11	26.500	152	25.000
1	25.700	102	24.100	120	26.500	118	25.000
4	25.600	142	24.100	114	26.500	124	25.000
114	25.600	124	24.000	17	26.500	156	24.700
17	25.500	7	23.600	1	26.400	116	23.200
140	25.400	156	23.600	19	26.400	113	23.000
19	25.400	146	23.200	18	26.400	148	21.600
33	25.400	116	22.490	33	26.400	146	21.600
36	25.320	113	21.800	153	26.370	38	19.900
154	25.300	148	21.700	30	26.300	40	16.000
151	25.300	38	19.600	112	26.200	151	15.800
22	25.300	40	16.000 (*)	154	26.100	138	2.630 (*)
18	25.300	138	2.520 (*)	133	26.100	149	1.064 (*)
126	25.200			13	26.030		

Table 30: Ratio of the measured to the calculated conductivity in synthetic precipitation samples (G1-G4).

Lab.No.	Measured value / calculated value				Remarks
	G1	G2	G3	G4	
1	0.90	0.96	0.96	0.97	
3	1.03	1.05	1.05	1.11	
4	1.00	1.00	1.01	1.01	
5	1.04	1.06	1.04	1.04	
6	0.93	0.95	0.97	0.97	
7	1.13	0.78	0.88	0.90	
8	1.03	1.02	1.01	1.01	
10	0.96	0.96	0.93	0.94	
11	1.08	1.11	1.15	1.12	
12	1.02	1.06	0.99	1.03	
13	1.04	1.00	0.98	0.97	
14	1.11	1.07	1.10	0.99	
15	1.03	1.05	1.06	1.04	
16	1.00	0.97	0.99	0.98	
17	0.74	0.79	0.87	0.90	
18	1.25	1.13	1.62	1.51	Reports only SO ₄ ²⁻ , NH ₄ ⁺ , Cl ⁻ , pH and cond.
19	1.28	1.25	1.22	1.25	
20	1.13	1.12	1.07	1.06	
21	1.05	1.05	1.04	1.04	
22	0.94	0.80	0.87	0.87	
23	0.97	1.03	1.01	0.99	
24					Conductivity is missing
26					Conductivity is missing
27	1.08	1.09	1.07	1.06	
30	0.96	0.98	1.01	1.03	
31	1.10	1.07	1.08	1.10	
32	0.98	0.95	1.00	0.96	
33	0.99	0.99	1.01	1.01	
34	1.07	1.04	1.10	1.10	
35	1.08	1.05	1.04	1.01	
36	1.08	1.05	1.01	1.03	
38	0.53	0.54	0.41	0.40	
39					Conductivity is missing SO ₄ ²⁻ is missing
40	0.95	1.05	1.12	1.20	
102	0.96	0.93	0.87	0.91	
104	1.16	1.22	1.21	1.24	
107	1.03	1.00	0.96	1.00	
108	1.00	0.89	0.94	1.00	
110	1.23	1.03	1.17	1.23	
111					
112	1.11	1.12	1.10	1.09	
113	0.91	0.92	1.07	0.99	
114	1.08	1.05	1.04	1.00	
115	1.21	1.14	1.11	1.07	
116	0.78	0.82	0.89	0.87	
117	1.61	1.43	1.54	1.23	
118	0.99	0.96	1.01	0.96	
120	1.02	1.03	1.00	1.00	
121	1.02	1.00	1.02	1.01	
124	0.98	0.99	0.96	1.00	
125					
126	1.06	1.00	1.01	1.02	
133	1.12	1.09	1.06	1.06	
137					
138	0.86	0.87	0.88	0.87	

Table 30, cont.

Lab.No.	Measured value / calculated value				Remarks
	G1	G2	G3	G4	
139	1.18	1.20	1.33	1.39	
140	1.09	1.07	1.05	1.07	
141					
142	0.96	0.93	0.87	0.91	
143					
144					
145	0.97	0.97	0.97	0.93	
146	0.98	1.00	1.05	0.90	
147	0.96	1.04	0.81	0.85	
148	0.74	0.78	0.98	0.93	
149	1.12	0.59	1.10	0.04	
151	1.19	1.18	1.17	0.71	
152	0.99	1.05	1.03	1.04	
153	1.02	1.05	1.14	1.11	
154	0.77	0.92	0.96	0.88	
156	1.52	1.61	1.06	1.13	
158	1.06	1.08	1.05	1.05	

Table 31: Ratio of equivalent concentration of anions to the equivalent concentration of cation measured in synthetic precipitation samples.

Lab.No.	Measured value / calculated value					Remarks
	G1	G2	G3	G4	Average	
1	0.93	0.98	0.96	0.98	0.96	
3	1.02	1.05	1.01	0.99	1.02	
4	1.02	1.02	0.98	0.97	1.00	
5	1.04	1.04	1.01	1.02	1.03	
6	0.99	0.99	0.98	0.98	0.98	
7	1.22	1.00	1.01	1.03	1.06	
8	1.09	1.07	1.05	1.05	1.07	
10	1.04	1.05	1.06	1.27	1.11	
11	1.23	1.19	1.33	1.27	1.26	
12	1.00	1.03	0.98	1.01	1.01	
13	1.07	1.09	1.09	1.08	1.08	
14	1.09	1.05	1.01	0.99	1.04	
15	1.03	1.07	1.00	1.02	1.03	
16	1.03	0.99	1.02	0.99	1.01	
17	0.87	0.93	1.03	1.08	0.97	
18	0.96	0.98	0.56	0.73	0.81	
19	1.23	1.19	1.16	1.19	1.19	
20	1.10	1.11	1.05	1.04	1.07	
21	1.06	1.06	1.02	1.02	1.04	
22	2.20	2.20	1.65	1.75	1.95	
23	1.04	1.06	1.03	1.00	1.03	
24	1.17	1.16	1.06	1.10	1.12	
26	1.06	1.09	1.01	1.02	1.05	
27	1.10	1.11	1.05	1.05	1.08	
30	1.16	1.15	0.95	0.93	1.05	
31	0.99	1.00	0.92	0.94	0.96	
32	1.04	1.06	1.00	1.03	1.03	
33	1.03	1.03	1.05	1.03	1.04	
34	1.11	1.04	1.01	1.06	1.06	
35	1.24	1.14	1.01	1.02	1.10	
36	1.05	1.04	0.98	1.00	1.02	
38	0.55	0.53	0.63	0.62	0.58	
39	1.08	1.04	0.99	1.01	1.03	
40	0.35	0.29	0.62	0.55	0.45	
102	1.01	0.94	0.97	0.93	0.96	
104	0.01	0.01	0.03	0.03	0.02	
107	1.05	1.04	0.85	1.00	0.98	
108	0.66	0.87	1.00	0.89	0.86	
109	1.05	1.05	0.92	0.95	0.99	
110	1.30	1.09	1.14	1.22	1.19	
112	1.05	1.05	0.92	0.95	0.99	
113	1.06	1.10	1.19	1.11	1.11	
114	1.07	1.02	1.01	0.99	1.03	
115	1.21	1.13	1.08	1.13	1.14	
116	1.13	1.15	1.14	1.22	1.16	
117	1.46	1.35	1.25	1.08	1.28	
118	1.15	1.14	1.06	1.07	1.11	
120	1.03	1.02	0.98	0.90	0.98	
121	0.96	0.97	0.96	0.98	0.97	
124	0.99	1.10	1.08	1.08	1.06	
126	0.89	0.97	0.96	0.97	0.95	
131	0.89	0.89	0.95	0.95	0.92	
133	1.41	1.38	1.51	1.38	1.42	
138	0.91	0.92	0.93	0.94	0.92	
139	1.18	1.12	1.09	1.10	1.12	

Table 31, cont.

Lab.No.	Measured value / calculated value					Remarks
	G1	G2	G3	G4	Average	
140	1.14	1.10	1.03	1.05	1.08	
142	1.01	0.94	0.97	0.93	0.96	
145	0.69	0.72	0.73	0.80	0.73	
146	1.04	1.07	1.05	1.05	1.05	
147	0.75	0.81	0.47	0.50	0.63	
148	1.33	1.32	1.21	1.17	1.26	
149	0.94	0.98	0.97	0.99	0.97	
151	0.72	0.73	0.73	0.71	0.72	
152	1.07	1.11	0.93	0.98	1.02	
153	1.20	1.13	1.02	1.04	1.10	
154	0.86	0.97	1.02	0.88	0.93	
156	1.89	1.39	1.54	1.45	1.57	
158	1.08	1.11	1.01	1.01	1.05	

Table 32: The ratio of the median values to the theoretical values for all parameters and samples.

Parameter	Sample No.	Median / Expected
SO ₄ -S	G1	0.99
	G2	0.99
	G3	1.00
	G4	0.99
NO ₃ -N	G1	0.97
	G2	0.98
	G3	0.98
	G4	0.99
NH ₄ -N	G1	1.00
	G2	1.02
	G3	1.00
	G4	1.03
pH (calc.From H ⁺)	G1	0.95
	G2	0.96
	G3	0.79
	G4	0.81
H	G1	0.82
	G2	0.96
	G3	0.75
	G4	0.49
Mg	G1	1.04
	G2	1.01
	G3	1.02
	G4	1.01
Na	G1	0.98
	G2	0.96
	G3	0.97
	G4	0.95
Cl	G1	0.97
	G2	0.98
	G3	0.96
	G4	0.94
Ca	G1	1.05
	G2	1.01
	G3	1.01
	G4	1.02
K	G1	0.98
	G2	0.95
	G3	0.98
	G4	0.98
Cond	G1	0.95
	G2	0.96
	G3	0.98
	G4	0.96

Table 33: Analytical methods used for the determination of chemical constituents in precipitation samples.

Constituents	Methods	Laboratory
SO ₄	1. Thorin	18
	2. Ion chromatography	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 31,32, 33, 34, 35, 36, 38, 104, 107,110, 111, 114, 115, 116, 118, 126, 130, 131, 133, 135, 136, 138, 140, 145, 146, 147, 151, 152, 154, 156, 158
	3. Capillary electrophoresis	39
	4. ICP-AES	109, 112, 113, 117
	5. FIA	121
NO ₃	1 Griess after Cd-red.	112,117
	2 Ion chromatography	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 31,32, 33, 34, 35, 36, 38, 107,110, 113, 114, 115, 116, 118, 126,130, 131, 133, 135, 136, 138, 140, 145,146,147,151,152,154,156,158
	3 UV-method/Photometric	40, 104,
	4 Capillary electrophoresis	39
	5 FIA	109, 111, 121
NH ₄	1 Indophenol	10, 19, 20, 32, 33, 34, 39, 40, 112, 114, 116, 117, 126, 140
	2 Berthelot reaction, salicylate	26, 118,
	3 Ion chromatography	1, 5, 6, 7, 8, 12, 13, 15, 21, 22, 23, 24, 31, 35, 36, 107,113, 115, 131,135,136,151,154,156,158
	4 Flow injection analysis (FIA)	14, 27, 109, 111, 121,145,146,147,152
	5 Chloramin T	16
	6 Nessler method	18,
	7 Kjeldahl	
	8 Photometry	3, 4, 104, 110,114, 133
H ⁺	Titration	6, 13,14, 31, 34, 104, 126,146, 154
Mg	1 Atomic absorption (AAS)	3, 4, 10, 16, 19, 22, 26, 27, 33, 34, 38, 39, 40, 116, 133, 154, 156
	2 Ion chromatography	1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 23, 31, 35, 36, 107, 113, 114, 126, 130, 131, 135, 136, 138, 140, 145, 146, 147,158
	3 ICP-AES	11, 104, 109, 111, 112, 115, 117, 118, 121,151,152
	4 ICP-MS	14
Na	1 AES	32, 33, 38, 39, 116, 133
	2 AAS	3, 4, 10, 16, 19, 26, 27, 34, 40,154,156
	3 ICP-AES	11, 104, 109, 110,111, 112, 115, 117,118, 121
	5 Ion chromatography	1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 22, 23, 31, 35, 36, 107, 113, 114, 126, 130, 131, 135, 136, 138, 140, 145, 146, 147, 151, 152, 158
	6 ICP-MS	14
Cl	1 Mercury thiocyanate-iron	18, 117, 40
	2 Ion chromatography	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 31,32, 33, 34, 35, 36, 38, 104, 107,110, 111, 113, 114, 115, 116, 118, 126, 130, 131, 133, 135, 136, 138, 140, 145, 146, 147, 151, 152, 154, 156, 158
	3 Capillary electrophoresis	39
	4 Potentiometric method	109, 112
	5 FIA	121
Ca	1 AAS	3, 4, 10, 16, 19, 22, 26, 27, 33, 34, 38, 39, 40, 116, 133, 154,156
	2 ICP-AES	11, 104, 109,110, 111, 112, 115,117, 118, 121, 151, 152
	3 Ion chromatography	1, 5, 6, 7, 8, 12, 15, 20,21, 23, 31, 35, 36, 107, 113, 114, 126, 130, 131, 135, 136, 138, 140, 145, 146, 147, 158
	6 AES	32,
	7 ICP-MS	14
K	1 AAS	3, 4, 10, 16, 19, 26, 27, 33, 34, 35, 40,154
	2 Ion chromatography	1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 22, 23, 31, 35, 36, 107, 113, 114, 126,130, 131, 135, 136, 138, 140, 145, 146, 147, 151, 158
	3 AES	32, 39, 116, 133, 156
	4 ICP-AES	11, 104, 109, 110,111, 112, 115, 117, 118, 121,152
	5 ICP-MS	14

Table 34: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples.

Lab. no.	SO ₄ ²⁻		NO ₃ ⁻		NH ₄ ⁺		Mg ²⁺		H ⁺ calc	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	0	0	0	-2	2	-7	2	-2	11	0
3	1	1	1	-1	2	10	2	-1	4	-14
4	0	2	1	-5	0	-1	2	0	3	-10
5	1	1	0	2	1	-1	1	-1	4	-10
6	1	-3	1	-2	2	-5	2	-6	4	0
7	2	4	0	1	1	-3	2	-5	15	-7
8	0	1	0	1	1	-3	0	0	2	-16
10	6	8	1	0	0	-5	1	-1	7	0
11	1	-2	1	-1	1	1	1	-6	2	-16
12	0	2	0	0	2	1	2	2	7	-4
13	0	5	5	3	6	-3	1	0	3	-13
14	1	4	1	-7	0	-2	1	-3	3	-19
15	1	-1	2	2	0	-4	1	-1	4	-13
16	2	-2	2	-2	3	3	2	-6	6	0
17	0	0	2	0	4	15	3	6	22	21
18	14	-32			25	-6			5	-13
19	7	-5	4	8	1	-2	1	0	4	-40
20	1	-2	1	0	3	-1	4	-14	2	-17
21	1	2	1	-2	0	0	0	0	4	-11
22	45	113	2	-3	3	16	1	-13	10	-18
23	0	-2	1	4	3	5	4	-5	5	-12
24	3	2	3	-2	2	-1	8	-2	2	-12
26	1	-2	2	-2	0	-1	1	1	2	-19
27	0	1	0	0	0	0	0	0	1	-19
30	4	10	4	4	26	21	26	15	10	-16
31	2	-10	2	5	10	9	5	-1	4	-15
32	3	-5	2	1	1	3			5	-5
33	1	4	0	-3	1	0	2	-4	5	-8
34	1	0	1	-4	4	-3	1	0	3	-18
35	2	5	0	-4	1	4	6	-42	3	-19
36	2	-3	0	-2	1	4	2	0	2	-17
38	2	-5	1	-2	2	0	0	-12	37	128
39	3	4	2	3	5	1	11	-4	3	-11
40	27	-99	5	26	23	185	10	-22	5	-58
102	3	3	17	1	6	4	3	8	8	2
104	18	-60	3	-1	3	-4	2	3	4	-17
107	3	6	1	-7	31	16	4	9	4	-14
108	22	-8	5	6	2	0	5	-74	14	14
110	1	-1	1	-3	2	-2	3	-28	13	-33
112	1	-8	4	-7	5	7	2	3	2	-23
113	3	-2	1	0	9	-8	7	17	6	-24
114	1	-3	0	1	2	2	6	-9	4	-14
115	1	-1	1	-3	11	-2	2	-9	5	-24
116	3	8	2	-2	1	6	14	-23	4	-9
117	3	-4	2	-6	1	-3	2	3	16	-57
118	5	11	1	-2	1	-2	2	1	3	-14
120	2	-6	1	2	4	12	1	-3	1	-13
121	4	-5	1	2	1	-1	1	-4	5	-10
124	2	1	6	3	8	9	6	-2	5	-15
125	3	3	2	4						
126	5	-2	4	-5	6	-12	5	11	4	-13
131	1	-1	2	1	5	-9	1	-1	15	19
133	7	16	23	0	8	-20	8	-42	3	-26
138	0	2	1	-1	2	-4	2	10	8	11
139	5	3	5	-4	2	2	0	2	5	-10
140	2	4	1	4	1	3	9	3	4	-21
142	3	3	17	1	6	4	3	8	8	2
145	9	-24	10	-22	3	2	3	7	6	0
146	1	-4	1	0	1	-4	1	0	5	-21
147	1	-4	2	-7	9	24	57	178	8	-27
148	2	17	1	0	1	-6	6	-24	4	-32
149	1	2	2	2	1	0	3	2	6	-5
151	15	-36	15	-31	8	-18	2	-5	2	-16
152	1	-2	2	-10	4	4	1	0	4	-20
153	2	-1	1	-6	2	-35	5	13	8	-25
154	5	5	3	4	11	31	26	169	19	2
156	4	1	6	-9	15	10	5	2	22	-69
158	0	-2	0	-2	3	7	3	-7		-19

Table 34, cont.

Lab. no.	Na ⁺		Cl ⁻		Ca ²⁺		K ⁺		Cond.	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	0	0	3	-7	2	2	1	-1	1	-1
3	5	1	1	-4	2	5	4	-1	2	0
4	1	10	4	-7	3	0	4	-2	2	-2
5	0	1	2	-5	0	2	3	1	0	3
6	2	0	2	-5	2	1	2	-4	4	-5
7	1	-1	1	-1	2	-6	1	-4	9	-5
8	1	-3	0	0	1	0	0	-2	2	-4
10	2	-8	24	26	3	1	3	5	0	0
11	1	-6	2	-5	1	-9	22	0	2	-1
12	1	-1	2	1	1	-6	2	0	2	5
13	4	-7	4	-7	2	-1	2	-3	1	-3
14	3	9	4	-21	1	1	8	13	3	-1
15	2	5	2	2	1	1	1	0	1	0
16	0	-2	4	-7	9	1	2	-7	2	-4
17	4	-17	18	63	3	-30	2	-57	3	-4
18			17	-26					3	-2
19	1	0			4	9	1	4	1	-2
20	3	-6	1	-6	5	-15	4	-3	0	1
21	2	-2	1	1	2	2	1	-3	0	1
22	8	-12	17	3	4	-31	15	-17	1	-1
23	3	-11	4	-2	7	-5	3	-3	1	0
24	2	-8	7	7	3	-5			0	0
26	0	1	2	2	1	3	1	2		
27	1	-2	0	-1	1	0	0	-3	1	-1
30	14	9	10	-3	13	14	31	2	4	-3
31	6	-1	4	-11	4	14	8	0	0	0
32	2	8	3	-7	12	-17	3	-4	2	-4
33	2	0	2	-1	2	-7	2	-5	1	-2
34	13	2	2	-3	2	7	4	-1	5	-2
35	13	9	2	-1	41	14	5	-18	1	-3
36	1	2	3	-4	1	5	2	-7	2	0
38	1	-2	3	-6	11	-3	1	-22	8	-9
39	2	-4	5	13	7	58	1	-3		
40	8	-4	6	6	26	-5	3	-2	9	-39
102	2	-7	20	-35	2	0	3	-4	2	-3
103					14	-103				
104	2	-5	13	-25	3	-6	4	-4	7	0
107	1	-2	4	-6	8	11	1	-3	0	-2
108	11	-9	12	-2	17	16	22	17	1	4
110	4	-7	3	2	8	-10	5	-2	3	-5
112	1	1			1	7	2	21	1	-2
113	5	-16	6	6	9	5	9	-7	6	-15
114	1	0	1	-2	6	10	2	-4	2	-1
115	5	-6	3	7	3	-8	16	-16	1	0
116	9	-13	2	-13	9	-71	2	-9	10	-17
117	1	-7	13	-2	1	3	2	3	1	-2
118	3	-2	5	-12	1	6	2	0	3	-6
120	7	-6	21	0	5	0	33	-8	3	-2
121	2	3			1	5	10	7	3	-1
122							30	-100		
124	6	-9	8	-9	6	-13	2	-6	3	-6
126	2	4	1	-5	24	21	3	-2	2	-1
131	0	-11	4	-5	3	-7	5	-14		
133	5	6	23	37	15	-18	11	-21	2	-4
138	4	-4	3	-6	1	7	3	1	2	-3
139	10	-32	4	8	1	-6	6	-19	1	20
140	6	2	4	2	12	13	8	-13	2	0
142	2	-7	20	-35	2	0	3	-4	2	-3
145	2	2	8	-9	7	19	5	-7	1	-5
146	1	2	2	-3	3	1	3	-2	3	-12
147	13	6	8	-22	136	469	2	-4	7	-7
148	8	43	16	-36	6	-64	4	30	16	-28
149	6	5	1	0	6	4	6	55	37	-22
151	3	-2	23	-36	1	-6	10	5	12	-3
152	3	-6	16	-12	2	5	2	-6	4	-9
153	2	-7	4	-2	7	25	11	52	12	-7
154	10	-17	16	14	11	-47	5	-30	3	-2
156	7	11	132	307	6	34	13	38	6	-8
158	2	5	2	-5	2	-7	4	-4	2	-4

Appendix 2

Figures




Figure 2: SO₂ in absorbing solution.




Figure 3: SO₂ in impregnated filter.




Figure 4: HNO₃ in impregnated filter.




Figure 5: NO₂ in absorbing solution.




Figure 6: NH₃ in impregnated filter.




Figure 7: Percent deviation from theoretical value for sulphate.




Figure 8: Percent deviation from theoretical value for nitrate.




Figure 9: Percent deviation from theoretical value for ammonium.




Figure 10: Percent deviation from theoretical value for pH.





Figure 11: Percent deviation from theoretical value for chloride.






Figure 13: Percent deviation from theoretical value for magnesium.




Figure 14: Percent deviation from theoretical value for calcium.




Figure 15: Percent deviation from theoretical value for potassium.




Figure 16: Percent deviation from theoretical value for conductivity.





Figure 17: The median compared to theoretical value.




Figure 18: Youden plot of SO₂-S in absorbing solution.




Figure 19: Youden plot of SO_2-S on impregnated filter.




Figure 20: Youden plot of $\text{HNO}_3\text{-N}$ on impregnated filter.





Figure 21: Youden plot of NO₂-N in absorbing solution.




Figure 22: Youden plot of NH₃-N on impregnated filter.




Figure 23: Youden plot of Ca in precipitation.





Figure 24: Youden plot of Cl in precipitation.




Figure 25: Youden plot of conductivity in precipitation.




Figure 26: Youden plot of K in precipitation.





Figure 27: Youden plot of Mg in precipitation.





Figure 28: Youden plot of Na in precipitation.




Figure 29: Youden plot of NH₄-N in precipitation.




Figure 30: Youden plot of strong acid in precipitation.




Figure 31: Youden plot of $\text{NO}_3\text{-N}$ in precipitation.




Figure 32: Youden plot of $\text{SO}_4\text{-S}$ in precipitation.




Figure 33: Youden plot of pH in precipitation.